



Airgas Mid America (LABORATORY)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 2-Jan-2013

Dear Sir,

This is your Certificate of Analysis:

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
1/2/2015	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Lot # AG300201

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.5 ppm	EB0010603	390.9 ppm
EB0010570	258.4 ppm	EB0010559	258.3 ppm
EB0010286	208.9 ppm	EB0010695	209.2 ppm
EB0010561	101.9 ppm	EB0010562	104.9 ppm
EB0010681	53.0 ppm	EB0010579	52.4 ppm

Analytical Method: NDIR

Analyst: 

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

AS IV Serial no: 026399
Version no: 7469

TEST RECORD 01685

Temp Date Time 210L

Air Blank: 12/04/14 20:28 .000

Calibration Check: 18 12/04/14 20:28 .104

Subject Name

Subject I.D.

Operator Name, I.D. *Menas*

Location *S01 Acauau*

220422

12-27-2014

AS IV Serial no: 026399
Version no: 7489

TEST RECORD 01695

Temp Date Time 210L

Air Blank: 12/04/14 20:31 .000

Calibration Check: 19 12/04/14 20:31 .103

Subject Name

Subject I.D.

Operator Name, I.D. *Menas*

Location *S01 Acauau*

220427

12-22-2014

AS IV Serial no: 026399
Version no: 7405

TEST RECORD 01607

Temp Date Time 210L

Air Blank: 12/04/14 20:35 .000

Calibration Check: 19 12/04/14 20:35 .103

Subject Name

Subject I.D.

Operator Name, I.D. *Menas*

Location *S01 Acauau*

220427

12-27-2014

AS IV Serial no: 026399
Version no: 7405

TEST RECORD 01686

Temp Date Time 210L

Void: RTI 12 12/04/14 20:37

Subject Name

Subject I.D.

Operator Name, I.D. *Menas*

Location *S01 Acauau*

220427

12-27-2014

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



BRAD M KERNS

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER; ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 12/27/2012

Number 220427

Expires 12/27/2014

Director of State Public Health Laboratory

ACTING DIRECTOR

Director, Department of Health