



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 9:42 am, Aug 06, 201
 REPORT#7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 030450	PRINTER SN 95.111.053	DATE OF INSPECTION 8-2-14
LOCATION OF INSTRUMENT (STREET AND CITY) 501 FARAON, ST. JOSEPH MO		TIME OF INSPECTION 2140

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (1 O-C - 40-C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS LOT # AG300201 EXP. DATE 01-02-2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 - .098	TEST 2 - .098	TEST 3 - .098
------------------	------------------	------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(.0-.04)	1	(.05-.09)	2	(.10-.14)	3	(.15-.19)	4	(OVER .19)	3
----------	---	----------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME ROBERT PAUL
TYPE II PERMIT NUMBER/EXPIRATION DATE 230306/ 12-11-2015	TELEPHONE NUMBER 816-271-4777

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



Airgas Mid America (LABORATORY)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 2-Jan-2013

Dear Sir,

This is your Certificate of Analysis:


<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
1/2/2015	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Lot # AG300201

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.5 ppm	EB0010603	390.9 ppm
EB0010570	258.4 ppm	EB0010569	258.3 ppm
EB0010285	208.9 ppm	EB0010595	209.2 ppm
EB0010561	101.9 ppm	EB0010562	104.9 ppm
EB0010681	53.0 ppm	EB0010579	52.4 ppm

Analytical Method: NDIR

Analyst: 

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

AS IV Serial no: 030450
Version no: 7410
TEST RECORD 05579 s/
Temp Date Time 210L
Air Blank: 08/02/14 21:44 .000
Calibration Check: 21 08/02/14 21:44 .058
Subject Name
Subject I.D.
Operator Name, I.D.
Robert Paul 5449
Location

AS IV Serial no: 030450
Version no: 7410
TEST RECORD 05600 s/
Temp Date Time 210L
Air Blank: 08/02/14 21:47 .000
Calibration Check: 23 08/02/14 21:47 .098
Subject Name
Subject I.D.
Operator Name, I.D.
Robert Paul 5449
Location

AS IV Serial no: 030450
Version no: 7410
TEST RECORD 05601 s/
Temp Date Time 210L
Air Blank: 08/02/14 21:50 .000
Calibration Check: 24 08/02/14 21:50 .098
Subject Name
Subject I.D.
Operator Name, I.D.
Robert Paul 5449
Location

AS IV Serial no: 030450
Version no: 7410
TEST RECORD 05602 s/
Temp Date Time 210L
Air Blank: 08/02/14 21:52
Calibration Check: 12 08/02/14 21:52
Subject Name
Subject I.D.
Operator Name, I.D.
Robert Paul 5449
Location



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

ROBERT L PAUL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/11/2013

NUMBER 230306

EXPIRES 12/11/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator PAUL, ROBERT
Permit No 230306
Date Issued 12/11/2013 Date Expires 12/11/2015