



Airgas Mid America (LABORATORY)
3600 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 2-Jan-2013

Dear Sir,

This is your Certificate of Analysis:


<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
1/2/2015	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Lot # AG300201

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010661	391.5 ppm	EB0010603	390.9 ppm
EB0010670	258.4 ppm	EB0010659	258.3 ppm
EB0010285	209.9 ppm	EB0010595	209.2 ppm
EB0010561	101.9 ppm	EB0010662	104.9 ppm
EB0010681	53.0 ppm	EB0010579	52.4 ppm

Analytical Method: NDIR

Analyst: 

ISO 17025:2005 A2LA accredited. Certificate Number 2909.01

AS IV Serial no: 030450
Version no: 7410
TEST RECORD 03055
Temp Date Time 210L s/
Void: RFI
12 12/04/14 15:10
Subject Name
Subject I.D.
Operator Name, I.D.
Location
220427
12-27-14

AS IV Serial no: 030450
Version no: 7410
TEST RECORD 03054
Temp Date Time 210L s/
Air Blank:
12/04/14 15:08 .000
Calibration Check:
26 12/04/14 15:08 .096
Subject Name
Subject I.D.
Operator Name, I.D.
Location
220427
12-27-14

AS IV Serial no: 030450
Version no: 7410
TEST RECORD 03053
Temp Date Time 210L s/
Air Blank:
12/04/14 15:05 .000
Calibration Check:
26 12/04/14 15:05 .096
Subject Name
Subject I.D.
Operator Name, I.D.
Location
220427
12-27-14

AS IV Serial no: 030450
Version no: 7410
TEST RECORD 03052
Temp Date Time 210L s/
Air Blank:
12/04/14 15:01 .000
Calibration Check:
25 12/04/14 15:01 .096
Subject Name
Subject I.D.
Operator Name, I.D.
Location
220427
12-27-14

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



BRAD M KERNS

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER; ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 12/27/2012

Number 220427

Expires 12/27/2014

Director of State Public Health Laboratory

ACTING DIRECTOR

Director, Department of Health