



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 3:52 pm, Aug 12, 2014

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 030788 PRINTER SN 84.9324.157 DATE OF INSPECTION 07/30/2014

LOCATION OF INSTRUMENT (STREET AND CITY) 400 First St Hillsboro TIME OF INSPECTION 2200hrs

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth lab LOT # 14100 EXP. DATE 05/01/2016

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34° SIMULATOR SN SD2002 SIMULATOR EXP DATE 03/30/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .102 TEST 2 .101 TEST 3

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE [Signature] 197 PRINT NAME Dr. Richard Boettje 197

TYPE II PERMIT NUMBER/EXPIRATION DATE 240268 06/12/2016 TELEPHONE NUMBER 636 1975800

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14110 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 5, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is May 1, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 030788
Version no: 532B

TEST RECORD 00031

Temp	Date	Time	s/ 210L
Air Blank:	07/30/14	22:05	.000
Calibration Check:	28 07/30/14	22:05	.102

Air Blank:
07/30/14 22:05 .000
Calibration Check:
28 07/30/14 22:05 .102

Subject Name

Test #1

Subject I.D.

Operator Name, I.D.

Dep R Beatta 197

Location

100 First St

N. Osborn, Mo 63050

AS IV Serial no: 030788
Version no: 532B

TEST RECORD 00033

Temp	Date	Time	s/ 210L
Air Blank:	07/30/14	22:09	.000
Calibration Check:	29 07/30/14	22:09	.101

Air Blank:
07/30/14 22:09 .000
Calibration Check:
29 07/30/14 22:09 .101

Subject Name

Test #3

Subject I.D.

Operator Name, I.D.

Dep Richard Beatta 197

Location

400 First St

N. Osborn, Mo 63050

AS IV Serial no: 030788
Version no: 532B

TEST RECORD 00032

Temp	Date	Time	s/ 210L
Air Blank:	07/30/14	22:07	.000
Calibration Check:	29 07/30/14	22:07	.101

Air Blank:
07/30/14 22:07 .000
Calibration Check:
29 07/30/14 22:07 .101

Subject Name

Test #2

AS IV Serial no: 030788
Version no: 532B

TEST RECORD 00034

Temp	Date	Time	s/ 210L
VOID: RFI	12 07/30/14	22:12	

VOID: RFI
12 07/30/14 22:12

Subject Name

RFI Test

Subject I.D.

400 First St

Hillsboro, Mo 63050

Hillsboro, Mo 63050

AS IV Serial no: 030788
Version no: 532B

TEST RECORD 00032

Temp Date Time 210L

9/

Air Blank:
07/30/14 22:07 .000
Calibration Check:
29 07/30/14 22:07 .101

Subject Name

Test #2

Subject I.D.

Operator Name, I.D.

Dep B Beatta 197

Location

400 First St

Hillsboro, Mo 63050

AS IV Serial no: 030788
Version no: 532B

TEST RECORD 00034

Temp Date Time 210L

9/

VOID: RFI
12 07/30/14 22:12

Subject Name

RFD Test

Subject I.D.

Operator Name, I.D.

Dep Richard Beatta 197

Location

400 First St

Hillsboro, Mo 63050



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
RICHARD BEATTIE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX EC/IR II, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/12/2014

NUMBER 240268

EXPIRES 6/12/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R5-10)

MO 560-0771 (6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator BEATTIE, RICHARD
 Permit No 240268
 Date issued 6/12/2014 Date Expires 6/12/2016