



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**

REPORT #7

By Carol Day at 11:10 am, Dec 08, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <b>030794</b>	PRINTER SN <b>84.9324.031</b>	DATE OF INSPECTION <b>12/7/2014</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>1781 Zumbetti Rd., St Charles, Mo 63303</b>		TIME OF INSPECTION <b>1706hrs</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) <b>ok</b>
<input checked="" type="checkbox"/> TEMPERATURE OF ALCO SENSOR (10°C - 40°C) <b>ok</b>
<input checked="" type="checkbox"/> PRINTER WORKING PROPERLY <b>ok</b>
<input checked="" type="checkbox"/> TIME AND DATE DISPLAYING PROPERLY <b>ok</b>

**BREATH ALCOHOL ACCURACY STANDARDS**

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <b>Guth Laboratories</b> LOT # <b>14030</b> EXP. DATE <b>1/20/2016</b>	
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <b>34.0°C</b> SIMULATOR SN <b>SD2248</b> SIMULATOR EXP DATE <b>1/17/2015</b>	
<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE	<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE	

TEST 1 • <b>.098</b>	TEST 2 • <b>.098</b>	TEST 3 • <b>.099</b>
----------------------	----------------------	----------------------

RFI DETECTOR OPERATING **ok**

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <b>5</b>	(.0-.04) <b>2</b>	(.05-.09) <b>3</b>	(.10-.14) <b>5</b>	(.15-.19) <b>8</b>	(OVER .19) <b>6</b>
-------------------	-------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE <b>Daniel Gibbons 324</b>	PRINT NAME <b>Daniel Gibbons 324</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>240034 2/11/2016</b>	TELEPHONE NUMBER <b>(636) 949-3300 x2324</b>

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901



## GUTH LABORATORIES, INC.

590 NORTH 87th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14030 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **January 22, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1215%** (w/vol) ethyl alcohol. The expiration date for this lot number is **January 20, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of  $0.100 \text{ g}/210\text{L} \pm 3\%$ .

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

AS IV Serial no: 030794  
Version no: 004C

TEST RECORD 00984

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
12/07/14 17:06 .000  
Calibration Check:  
22 12/07/14 17:06 .098

Subject Name

Subject I.D.

Operator Name, I.D.

D. Gibbons 324

Location

1781 Zumbetti Rd.

St Charles, MO 63303

AS IV Serial no: 030794  
Version no: 004C

TEST RECORD 00985

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
12/07/14 17:07 .000  
Calibration Check:  
23 12/07/14 17:07 .098

Subject Name

Subject I.D.

Operator Name, I.D.

D. Gibbons 324

Location

1781 Zumbetti Rd.

St Charles, MO 63303

AS IV Serial no: 030794  
Version no: 004C

TEST RECORD 00986

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
12/07/14 17:09 .000  
Calibration Check:  
23 12/07/14 17:09 .099

Subject Name

Subject I.D.

Operator Name, I.D.

D. Gibbons 324

Location

1781 Zumbetti Rd.

St Charles, MO 63303

AS IV Serial no: 030794  
Version no: 004C

TEST RECORD 00987

Temp Date Time <sup>s/</sup> 210L

Void: RFI  
12 12/07/14 17:10

Subject Name

Subject I.D.

Operator Name, I.D.

D. Gibbons 324

Location

1781 Zumbetti Rd.

St Charles, MO 63303