



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 1:12 pm, Sep 08, 2014

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <u>030800</u>	PRINTER SN <u>84.9324.155</u>	DATE OF INSPECTION <u>08/27/14</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>400 First St Hillsboro</u>		TIME OF INSPECTION <u>0040 hrs</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Lab LOT # 14110 EXP. DATE 05/01/2016

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34° SIMULATOR SN SD2222 SIMULATOR EXP DATE 03/21/15

CALIBRATION CHECK -- (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

~~0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE~~

~~0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE~~

TEST 1 = .101 TEST 2 = .101 TEST 3 = .101

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument functions properly.

INSPECTING OFFICER

SIGNATURE <u>Dep Heather 197</u>	PRINT NAME <u>Dep Richard Beattie 197</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>240268 06/12/2016</u>	TELEPHONE NUMBER <u>636 797-5000</u>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14110 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 5, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is May 1, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS 10 Serial no: 0300000
Version no: 532B

TEST RECORD 000036

Temp Date Time 210L

Air Blank: 08/28/14 00:46 .000

Calibration Check: 25 08/28/14 00:46 .101

Subject Name

test #3

Subject I.D.

Operator Name, I.D.

D.P. Bosthe 197

Location

400 First St

Albion, Mo

AS 10 Serial no: 0300000
Version no: 532B

TEST RECORD 000037

Temp Date Time 210L

VOID: RFI

12 08/28/14 00:47

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

D.P. Bosthe 197

Location

400 First St

Albion, Mo

AS 10 Serial no: 8308808
Version no: 5322

TEST RECORD 000935

Temp Date Time 219L

Air Blank:

08/28/14 00:44 .000

Calibration Check:

24 08/28/14 00:44 .101

Subject Name

Test #2

Subject I.D.

Operator Name: I.D.

408 Seattle 197

Location

408 First St

Willsboro, Mo

AS 10 Serial no: 8308808
Version no: 5322

TEST RECORD 000934

Temp Date Time 219L

Air Blank:

08/28/14 00:42 .000

Calibration Check:

23 08/28/14 00:42 .101

Subject Name

Test #1

Subject I.D.

Operator Name: I.D.

408 Seattle 197

Location

408 First St

Willsboro, Mo



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

RICHARD BEATTIE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX EC/IR II, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/12/2014

NUMBER 240268

EXPIRES 6/12/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R5-10)

MD 580-0771 (6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator BEATTIE, RICHARD
Permit No 240268
Date Issued 6/12/2014 Date Expires 6/12/2016