



AS IV Serial no: 030805  
Version no: 532B

TEST RECORD 00094

Temp Date Time 210L

VOID: RFI  
12 06/17/14 00:17

Subject Name

RTI TEST

Subject I.D.

LANCE 240205

Operator Name, I.D.

Location

302 S. MYND ST

See No 64863

AS IV Serial no: 030805  
Version no: 532B

TEST RECORD 00093

Temp Date Time 210L

Air Blank: 06/17/14 00:16 .000  
Calibration Check: 25 06/17/14 00:16 .100

Subject Name

TEST # 3

Subject I.D.

LANCE 240205

Operator Name, I.D.

Location

302 S. MYND ST

See No 64863

AS IV Serial no: 030805  
Version no: 532B

TEST RECORD 00092

Temp Date Time 210L

Air Blank: 06/17/14 00:14 .000  
Calibration Check: 24 06/17/14 00:14 .099

Subject Name

TEST # 2

Subject I.D.

LANCE 240205

Operator Name, I.D.

Location

302 S. MYND ST

See no 64863

AS IV Serial no: 030805  
Version no: 532B

TEST RECORD 00091

Temp Date Time 210L

Air Blank: 06/17/14 00:13 .000  
Calibration Check: 23 06/17/14 00:13 .100

Subject Name

TEST # 1

Subject I.D.

LANCE 240205

Operator Name, I.D.

Location

302 S. MYND ST

See No 64863



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470



## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13010** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **January 14, 2013**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1218%** (w/vol) ethyl alcohol. The expiration date for this lot number is **January 9, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*