MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 3:52 pm, Aug 12, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
03A.0436.100	84.9324.142	07/30/2014
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
400 First St. Hillsboro		2215 hrs

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- [x] TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- [x] PRINTER WORKING PROPERLY
- [x] TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- [x] SIMULATOR SOLUTION
- [ ] COMPRESSED ETHANOL-GAS MIXTURE
- [x] STANDARD SUPPLIER Guth Lab LOT # 14110 EXP. DATE 05/01/2016
- [x] SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34° SIMULATOR SN SD2220 SIMULATOR EXP DATE 03/20/2015
- [x] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
  Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - [ ] ~~0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE~~
  - [ ] ~~0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE~~

TEST 1	TEST 2	TEST 3
.099	.099	.098

- [x] RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE	PRINT NAME
[signature] 197	[signature] 197
TYPE/PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
240268 06/12/2016	636 797-5000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14110** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 5, 2014,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 1, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
*Testing was conducted using Cerilliant Reference Standard lot number **FN122211-02** whose values are traceable to NIST.*
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

S IV Serial no: 030807
ersion no: 532B

TEST RECORD 00030

emp	Date	Time	9/ 210L
ir Blank:			
	07/30/14	22:30	.000
alibration Check:			
26	07/30/14	22:30	.099

ubject Name
Test #0

ubject I.D.

perator Name, I.D.
Dep R Beattie 197

ocation
400 First St
Hillsboro, Mo 63050

AS IV Serial no: 030807
Version no: 532B

TEST RECORD 00029

Temp	Date	Time	9/ 210L
Air Blank:			
	07/30/14	22:28	.000
Calibration Check:			
26	07/30/14	22:28	.099

Subject Name
Test #1

Subject I.D.

Operator Name, I.D.
Dep R Beattie 197

Location
400 First St
Hillsboro, Mo 63050

IV Serial no: 030807
sion no: 532B

TEST RECORD 00031

mp	Date	Time	9/ 210L
r Blank:			
	07/30/14	22:32	.000
libration Check:			
6	07/30/14	22:32	.099

bject Name
Test #3

AS IV Serial no: 030807
Version no: 532B

TEST RECORD 00032

Temp	Date	Time	9/ 210L
VOID: RFI			
12	07/30/14	22:34	

Subject Name
Dep R Beattie 197

AS IV Serial no: 030807
Version no: 532B

TEST RECORD 00031

Temp	Date	Time	9/ 210L

Air Blank:
07/30/14 22:32 .000
Calibration Check:
26 07/30/14 22:32 .099

Subject Name
Test 49

Subject I.D.
Dep R Beatte 197

Operator Name, I.D.

Location
400 First St
Hillsboro, Mo 63050

AS IV Serial no: 030807
Version no: 532B

TEST RECORD 00032

Temp	Date	Time	9/ 210L

VOID: RFI
12 07/30/14 22:34

Subject Name
Dep R Beatte 197

Subject I.D.
RFI Test

Operator Name, I.D.
Dep R Beatte 197

Location
400 First St
Hillsboro, Mo 63050

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

RICHARD BEATTIE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX EC/IR II, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/12/2014

NUMBER 240268

EXPIRES 6/12/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R5-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BEATTIE, RICHARD
Permit No 240268
Date Issued 6/12/2014 **Date Expires** 6/12/2016