



AS 10 Serial no: 000000  
Version: 001 0000

AS 10 Serial no: 000000  
Version: 001 0000

AS 10 Serial no: 000000  
Version: 001 0000

TEST RECORD 00920

TEST RECORD 00929

TEST RECORD 00928

Temp Date Time 210L

Temp Date Time 210L

Temp Date Time 210L

AS 10 SPI

AS 10 SPI

AS 10 SPI

12 01/22/14 09:52

01/22/14 09:50 .000

01/22/14 09:47 .000

Subject Name

Calibration Check

Calibration Check

RPI Test

Test # 3

Test # 2

Test # 1

Subject I.D.  
W.L. Davenport 008

Subject I.D.  
W.L. Davenport 008

Subject I.D.  
W.L. Davenport

Subject I.D.  
W.L. Davenport

Operator Name  
MCSO

Operator Name, I.D.  
MCSO

Operator Name, I.D.  
MCSO

Operator Name, I.D.  
MCSO

Location

Location

Location

State of Missouri  
DEPARTMENT OF HEALTH



P E R M I T  
TYPE II



WILLIAM DAVENPORT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER; ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/28/2012

Number 220309

Expires 09/28/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)



**GUTH LABORATORIES, INC.**

590 NORTH 87th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13010 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 14, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1218% (w/vol) ethyl alcohol. The expiration date for this lot number is January 9, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*