



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED REPORT #
 By Carol Day at 3:55 pm, May 05, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|-----------------------------|---------------------------|----------------------------------|
| ALCO SENSOR IV SN 030809 | PRINTER SN 84.9324.042 | DATE OF INSPECTION 04/28/2014 |
|-----------------------------|---------------------------|----------------------------------|

| | |
|---|-------------------------------|
| LOCATION OF INSTRUMENT (STREET AND CITY) 2410 Goodale Overland | TIME OF INSPECTION 5:30 pm |
|---|-------------------------------|

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth Laboratories LOT # 14030 EXP. DATE 01/20/2016
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2738 SIMULATOR EXP DATE 07/10/2014

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

| | | |
|---------------|---------------|---------------|
| TEST 1 → .095 | TEST 2 → .095 | TEST 3 → .096 |
|---------------|---------------|---------------|

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | | | | | | | |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 0 | (0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 0 | (OVER .19) | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Software chip update to version 532b. Lithium battery for time keeping replaced (Intoximeter). Wire holding clock board broken. Repaired (Intoximeter)

| | |
|--|------------------------------------|
| INSPECTING OFFICER | |
| SIGNATURE P.O. Thomas M. Moore #554 | PRINT NAME P.O. Thomas M. Moore |
| TYPE II PERMIT NUMBER/EXPIRATION DATE 230293 11/26/2015 | TELEPHONE NUMBER (314) 428-1221 |

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 030809
Version no: 532B

TEST RECORD 00038
Temp Date Time 210L
a/

VOID: RFI
12 04/28/14 17:40
Subject Name

MAINTENANCE
Subject I.D.

Operator Name, I.D.

MOORE 554
Location
2410 GRADUATE

AS IV Serial no: 030809
Version no: 532B

TEST RECORD 00037
Temp Date Time 210L
a/

Air Blank:
04/28/14 17:35 .000
Subject Test: Auto
26 04/28/14 17:35 .096

Subject Name
MAINTENANCE
Subject I.D.

Operator Name, I.D.

MOORE 554
Location
2410 GRADUATE

AS IV Serial no: 030809
Version no: 532B

TEST RECORD 00036
Temp Date Time 210L
a/

Air Blank:
04/28/14 17:32 .000
Subject Test: Auto
25 04/28/14 17:32 .095

Subject Name
MAINTENANCE
Subject I.D.

Operator Name, I.D.

MOORE #554
Location
2410 GRADUATE

AS IV Serial no: 030809
Version no: 532B

TEST RECORD 00035
Temp Date Time 210L
a/

Air Blank:
04/28/14 17:30 .000
Subject Test: Auto
25 04/28/14 17:30 .095

Subject Name
MAINTENANCE
Subject I.D.

Operator Name, I.D.

MOORE 554
Location
2410 GRADUATE



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14030** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **January 22, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1215%** (w/vol) ethyl alcohol. The expiration date for this lot number is **January 20, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** $\pm 3\%$.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
THOMAS M MOORE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX EC/IR II, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/26/2013

NUMBER 230293

EXPIRES 11/26/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MOORE, THOMAS
Permit No 230293
Date Issued 11/26/2013 Date Expires 11/26/2015