



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**

By Carol Day at 3:39 pm, Apr 28, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <b>043579</b>	PRINTER SN <b>91.9821.022</b>	DATE OF INSPECTION <b>4-27-14</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>101 Sher. of Dierker Ct. O'Fallon</b>		TIME OF INSPECTION <b>2353</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) <b>passed</b>
<input checked="" type="checkbox"/> TEMPERATURE OF ALCO SENSOR (10°C - 40°C) <b>25°C</b>
<input checked="" type="checkbox"/> PRINTER WORKING PROPERLY <b>passed</b>
<input checked="" type="checkbox"/> TIME AND DATE DISPLAYING PROPERLY <b>2356 4-27-14</b>

**BREATH ALCOHOL ACCURACY STANDARDS**

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <b>Intoximeter</b> LOT # <b>#G402703</b> EXP. DATE <b>1-27-16</b>	
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____	

**CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
<input checked="" type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • <b>.077</b>	TEST 2 • <b>.077</b>	TEST 3 • <b>.077</b>
----------------------	----------------------	----------------------

RFI DETECTOR OPERATING **passed**

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <b>0</b>	(0-.04) <b>0</b>	(.05-.09) <b>0</b>	(.10-.14) <b>0</b>	(.15-.19) <b>0</b>	(OVER .19) <b>0</b>
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**Tru Cal .077**  
  
**Software Upgrade 532 B**

**INSPECTING OFFICER**

SIGNATURE <b>Michael C Hoekle</b>	PRINT NAME <b>Michael C Hoekle</b>
TYPE (PERMITS NUMBER/EXPIRATION DATE) <b>220180 8-3-14</b>	TELEPHONE NUMBER <b>636-949-0809</b>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

Customer Name  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

Test Date: 28-Jan-2014

Lot # AG402703

<u>Exp. Date</u> 27-Jan-2016	<u>Cyl. Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0,080 ± 0.002 BrAC (218 ppm) Balance
---------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control  
 Date: 2014.01.28 12:39:44 -08:00  
 Reason: Dry gas standard certification of analysts  
 Location: Airgas USA LLC (Lab)

Analyst:   
 Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 2989.01**

AS 13 Serial no: 843579  
Version no: 5225

TEST REPORT 00017

Temp Date Time 210L

AV Blank

04/27/14 23:55 .000

Callibration Check

25 04/27/14 23:55 .077

Subject Name

Subject I.D.

Operator Name: I.D.

focke 570

Location

101 Sten P Dierker G

AS 14 Serial no: 843579  
Version no: 5225

TEST REPORT 00018

Temp Date Time 210L

AV Blank

04/27/14 23:58 .000

Callibration Check

25 04/27/14 23:58 .077

Subject Name

Subject I.D.

Operator Name: I.D.

focke 570

Location

101 Sten P Dierker G

AS 15 Serial no: 843579  
Version no: 5225

TEST REPORT 00019

Temp Date Time 210L

AV Blank

04/28/14 09:59 .000

Callibration Check

25 04/28/14 09:59 .077

Subject Name

Subject I.D.

Operator Name: I.D.

focke 570

Location

101 Sten P Dierker G

AS 16 Serial no: 843579  
Version no: 5225

TEST REPORT 00020

Temp Date Time 210L

AV Blank

12 04/28/14 08:01

Subject Name

Subject I.D.

Operator Name: I.D.

focke 570

Location

101 Sten P Dierker G

State of Missouri  
DEPARTMENT OF HEALTH



P E R M I T  
TYPE II



MICHAEL C HOEFLE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 08/03/2012

Number 220180

Expires 08/03/2014

Director of State Public Health Laboratory

Director, Department of Health

MO 580-0771 (7-88)

Lab. 4 (R7-88)