



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check. Send copy to Department of Health and Senior Services; retain original in department file.

By Carol Day at 9:32 am, May 27, 2014

| | | |
|--|----------------------------|----------------------------------|
| ALCO SENSOR IV SN 062093 | PRINTER SN 093.3563.014 | DATE OF INSPECTION 05/25/2014 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 1001 Heroes Way, Sugar Creek, MO | | TIME OF INSPECTION 3:50 pm |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Gulth Laboratories LOT # 14030 EXP. DATE 01/20/2016

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD3136 SIMULATOR EXP DATE 07/15/2014

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 \approx .100 TEST 2 \approx .099 TEST 3 \approx .100

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | | | | | | | |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 0 | (0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 1 | (.15-.19) | 0 | (OVER .19) | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

| | |
|---|------------------------------------|
| INSPECTING OFFICER | |
| SIGNATURE | PRINT NAME Wesley H. Edwards |
| TYPE II PERMIT NUMBER/EXPIRATION DATE 230200/ 09/19/2015 | TELEPHONE NUMBER (816) 252-5560 |

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 062093
Version no: 532B

TEST RECORD 00234

Temp Date Time ^{s/} 210L

Air Blank:

05/25/14 15:55 .000

Calibration Check:

23 05/25/14 15:55 .100

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

Edwards 908

Location

1001 Heroes Way

Sugar Creek, MO

AS IV Serial no: 062093
Version no: 532B

TEST RECORD 00232

Temp Date Time ^{s/} 210L

Air Blank:

05/25/14 15:50 .000

Calibration Check:

22 05/25/14 15:50 .100

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

Edwards 908

Location

1001 Heroes Way

Sugar Creek, MO

AS IV Serial no: 062093
Version no: 532B

TEST RECORD 00235

Temp Date Time ^{s/} 210L

VOID: RFI

12 05/25/14 15:56

Subject Name

Test 4

Subject I.D.

Operator Name, I.D.

Edwards 908

Location

1001 Heroes Way

Sugar Creek, MO

AS IV Serial no: 062093
Version no: 532B

TEST RECORD 00233

Temp Date Time ^{s/} 210L

Air Blank:

05/25/14 15:52 .000

Calibration Check:

23 05/25/14 15:52 .099

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

Edwards 908

Location

1001 Heroes Way

Sugar Creek, MO



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14030 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 22, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1215% (w/vol) ethyl alcohol. The expiration date for this lot number is January 20, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

WESLEY H EDWARDS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE: 9/19/2013

NUMBER 230200

EXPIRES 9/19/2015

Wesley H Edwards

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Garl Vesterby

acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 590.0271 (8/10)

LAB-1 (7/9/10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator EDWARDS, WESLEY
Permit No 230200
Date issued 9/19/2013 Date Expires 9/19/2015