





STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**BRETT DAVIS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOXILYZER 5000, INTOXILYZER 8000, ALCO-SENSOR IV  
 W/PRINTER**


for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/14/2013

NUMBER 230157

EXPIRES 8/14/2015

  
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

  
 acting director  
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO SR0-0771 (6-10)

LAB-4 (PE-10)

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator DAVIS, BRETT  
 Permit No 230157  
 Date Issued 8/14/2013 Date Expires 8/14/2015



## GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-664-6470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13210** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **July 31, 2013**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1216%** (w/vol) ethyl alcohol. The expiration date for this lot number is **July 29, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

AS IV Serial no: 087967  
Version no: 004C

TEST RECORD 00810

Temp Date Time 210L  
s/

Air Blank:  
01/23/14 22:52 .000  
Subject Test: Auto  
23 01/23/14 22:52 .100

Subject Name

Subject I.D.

Operator Name, I.D.

*B DB 9004*  
Location

*Joplin Jail*

*0000*

AS IV Serial no: 087967  
Version no: 004C

TEST RECORD 00812

Temp Date Time 210L  
s/

Air Blank:  
01/23/14 22:56 .000  
Subject Test: Auto  
24 01/23/14 22:56 .099

Subject Name

Subject I.D.

Operator Name, I.D.

*B DB 9004*  
Location

*Joplin Jail*

AS IV Serial no: 087967  
Version no: 004C

TEST RECORD 00811

Temp Date Time 210L  
s/

Air Blank:  
01/23/14 22:54 .000  
Subject Test: Auto  
24 01/23/14 22:54 .099

Subject Name

Subject I.D.

Operator Name, I.D.

*B DB 9004*  
Location

*Joplin Jail*

AS IV Serial no: 087967  
Version no: 004C

TEST RECORD 00813

Temp Date Time 210L  
s/

Void: RFI  
12 01/23/14 22:57

Subject Name

Subject I.D.

Operator Name, I.D.

*B DB 9004*  
Location

*Joplin Jail*