



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

received 3/1/14-cd

REPORT #7

**REVIEWED**

By Carol Day at 2:11 pm, Mar 14, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check. Send copy to Department of Health and Senior Services; retain original in department file.

|  |                            |                                  |
|--|----------------------------|----------------------------------|
| ALCO SENSOR IV SN<br>087967  | PRINTER SN<br>08C.3556.200 | DATE OF INSPECTION<br>02/24/2014 |
| LOCATION OF INSTRUMENT (STREET AND CITY)<br>303 E. 3rd Street Joplin (DWI-2 Unit 2119) |                            | TIME OF INSPECTION<br>2:47 pm    |

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH LAB LOT # 13210 EXP. DATE 07/29/2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD3001 SIMULATOR EXP DATE 07/10/2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .098

TEST 2 • .099

TEST 3 • .099

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

|          |   |         |   |           |   |           |   |           |   |            |   |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 0 | (0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 0 | (OVER .19) | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**


SIGNATURE

PRINT NAME  
Brett Davis

TYPE II PERMIT NUMBER/EXPIRATION DATE  
230157/ 08-14-2015

TELEPHONE NUMBER  
(417) 623-3131

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**BRETT DAVIS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOXILYZER 5000, INTOXILYZER 8000, ALCO-SENSOR IV  
W/PRINTER**


for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE: 8/14/2013

NUMBER: 230157

EXPIRES: 8/14/2015

  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

  
acting director  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MQ 680-8771 (6-10)

LAB-4 (02-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator: DAVIS, BRETT  
Permit No: 230157  
Date issued 8/14/2013 Date Expires 8/14/2015



# GUTH LABORATORIES, INC.

500 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13210 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1216% (w/vol) ethyl alcohol. The expiration date for this lot number is July 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

**NIST Traceability:**

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 087967  
Version no: 004C

TEST RECORD 00836

Temp Date Time 210L

Air Blank:  
02/24/14 14:50 .000  
Subject Test: Auto  
28 02/24/14 14:50 .099

Subject Name

Subject I.D.

Operator Name, I.D.

*B DS 9004*

Location

*Joplin Jail*

AS IV Serial no: 087967  
Version no: 004C

TEST RECORD 00837

Temp Date Time 210L

Void: REL  
12 02/24/14 14:52

Subject Name

Subject I.D.

Operator Name, I.D.

*B DS 9004*

Location

*Joplin Jail*

AS IV Serial no: 087967  
Version no: 004C

TEST RECORD 00834

Temp Date Time 210L

Air Blank:  
02/24/14 14:47 .000  
Subject Test: Auto  
27 02/24/14 14:47 .038

Subject Name

Subject I.D.

Operator Name, I.D.

*B DS 9004*

Location

*Joplin Jail*

AS IV Serial no: 087967  
Version no: 004C

TEST RECORD 00835

Temp Date Time 210L

Air Blank:  
02/24/14 14:48 .000  
Subject Test: Auto  
28 02/24/14 14:48 .099

Subject Name

Subject I.D.

Operator Name, I.D.

*B DS 9004*

Location

*Joplin Jail*