



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 8:55 am, Nov 04, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 087967	PRINTER SN 08C.3556.200	DATE OF INSPECTION 10/31/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 303 E. 3rd Street Joplin		TIME OF INSPECTION 11:43 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters, Inc. LOT # AG319103 EXP. DATE 03/10/2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .102

TEST 2 ➡ .102

TEST 3 ➡ .102

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	1	(.10-.14)	4	(.15-.19)	1	(OVER .19)	2
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE


PRINT NAME
Brett Davis

TYPE / PERMIT NUMBER / EXPIRATION DATE
 230157 / 08-14-2015

TELEPHONE NUMBER
 (417) 623-3131

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
BRETT DAVIS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOXILYZER 5000, INTOXILYZER 8000, ALCO-SENSOR IV
W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 572.020 through 572.041, RSMo and 308.111 through 308.119 RSMo.

DATE: 8/14/2013
NUMBER: 230157
EXPIRES: 8/14/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (8-10) LAB-100-101

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The name of operator is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath (exp. of expired air) in Missouri.

Operator: DAVIS, BRETT
Permit No: 230157
Date Issued: 8/14/2013 Date Expires: 8/14/2015



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Inoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 11-Jul-2013

Lot # AG319103

Exp. Date
10-Mar-2015

Cyl. Type
30

Component
Ethanol
Nitrogen

Certified Concentration
0-100 ± 2% BrAC (272 ppm)
Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2013.07.11:10:33:59 -06:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Analyst: _____

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

AS IV Serial no: 087967
Version no: 532B

TEST RECORD 00162

Temp Date Time ^{s/} 210L

Air Blank:
10/31/14 23:43 .000
Calibration Check:
21 10/31/14 23:43 .102

Subject Name

Subject I.D.

Operator Name, I.D.

Location

Joplin Jail

AS IV Serial no: 087967
Version no: 532B

TEST RECORD 00164

Temp Date Time ^{s/} 210L

Air Blank:
10/31/14 23:47 .000
Calibration Check:
21 10/31/14 23:47 .102

Subject Name

Subject I.D.

Operator Name, I.D.

Location

Joplin Jail

AS IV Serial no: 087967
Version no: 532B

TEST RECORD 00163

Temp Date Time ^{s/} 210L

Air Blank:
10/31/14 23:45 .000
Calibration Check:
21 10/31/14 23:45 .102

Subject Name

Subject I.D.

Operator Name, I.D.

Location

Joplin Jail

AS IV Serial no: 087967
Version no: 532B

TEST RECORD 00165

Temp Date Time ^{s/} 210L

VOID: RFI
12 10/31/14 23:48

Subject Name

Subject I.D.

Operator Name, I.D.

Location

Joplin Jail