



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

RECEIVED 1/14/14 **REPORT #7**

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

**REVIEWED**  
By Carol Day at 9:21 am, Mar 14, 2014

ALCO SENSOR IV SN <u>087971</u>	PRINTER SN <u>08C, 3556.195</u>	DATE OF INSPECTION <u>1/8/14</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>330 N 2nd St Poplar Bluff MO 63901</u>		TIME OF INSPECTION <u>0930</u>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C) 17°C

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH LAB LOT # 13210 EXP. DATE 7/29/15

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34°C SIMULATOR SN SD 2748 SIMULATOR EXP DATE 1/10/14

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 <u>.098</u>	TEST 2 <u>.096</u>	TEST 3 <u>.097</u>
--------------------	--------------------	--------------------

RFI DETECTOR OPERATING PASSED

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <u>0</u>	(0-.04) <u>0</u>	(.05-.09) <u>0</u>	(.10-.14) <u>0</u>	(.15-.19) <u>0</u>	(OVER .19) <u>0</u>
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

No New Parts Added

**INSPECTING OFFICER**

SIGNATURE <u>CM Richard W. Kuma #255</u>	PRINT NAME <u>RICHARD W. KUMA</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>230135 7/11/15</u>	TELEPHONE NUMBER <u>573 785 5776</u>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901



**GUTH LABORATORIES, INC.**

80 NORTH 6TH STREET • HARRISBURG, PA 17111-4911 • TELEPHONE: 717-634-4470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13210 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1216% (w/vol) ethyl alcohol. The expiration date for this lot number is July 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

AS IV Serial no: 087971  
Version no: 004C

TEST RECORD 00339

Temp Date Time 210L

Air Blank: 01/08/14 09:35 .000

Subject Test: Man

20 01/08/14 09:35 .097

Subject Name

Subject I.D.

Operator Name: I.D.

Location

AS IV Serial no: 087971  
Version no: 004C

TEST RECORD 00337

Temp Date Time 210L

Air Blank: 01/08/14 09:30 .000

Subject Test: Man

18 01/08/14 09:30 .098

Subject Name

Subject I.D.

Operator Name: I.D.

Location

AS IV Serial no: 087971  
Version no: 004C

TEST RECORD 00338

Temp Date Time 210L

Air Blank: 01/08/14 09:32 .000

Subject Test: Man

19 01/08/14 09:32 .096

Subject Name

Subject I.D.

Operator Name: I.D.

Location

AS IV Serial no: 087971  
Version no: 004C

TEST RECORD 00340

Temp Date Time 210L

Void/ RFI  
12 01/08/14 09:37

Subject Name

Subject I.D.

Operator Name: I.D.

Location



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT  
TYPE II**

**RICHARD W KNAPP**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 5000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/11/2013

NUMBER 230135

EXPIRES 7/11/2015

MO 660-0771 (8-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator **KNAPP, RICHARD**  
Permit No **230135**  
Date Issued **7/11/2013** Date Expires **7/11/2015**