



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

received 1/28/14 REPORT # 2 d

**REVIEWED**  
 Instrument is repaired  
 By Carol Day at 1:08 pm, Feb 04, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and when needed. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 087972	PRINTER SN 08C.3556.386	DATE OF INSPECTION 01/19/2014
LOCATION OF INSTRUMENT (STREET AND CITY) TROOP D, ZONE 3 ZONE OFFICE, BOLIVAR		TIME OF INSPECTION 11:23 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C) 20°C
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER REPCO MARKETING INC LOT # 13001 EXP. DATE 03/07/2015
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.1 SIMULATOR SN G6804 SIMULATOR EXP DATE 07/15/2014

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 <u>.105</u>	TEST 2 <u>.104</u>	TEST 3 <u>.105</u>
--------------------	--------------------	--------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	1	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

<b>INSPECTING OFFICER</b>	PRINT NAME
SIGNATURE <u>Joshua L White</u>	JOSHUA L. WHITE
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
230208 / 09-23-2015	(417) 895-6868

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT  
TYPE II**

**JOSHUA L WHITE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/23/2013

NUMBER 230208

EXPIRES 9/23/2015

  
\_\_\_\_\_  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

  
\_\_\_\_\_  
acting director  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680-0771 (8-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator WHITE, JOSHUA  
Permit No 230208  
Date Issued 9/23/2013 Date Expires 9/23/2015

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 13001**

**EXPIRATION DATE: March 7, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013  
The expiration date for this lot number is March 7, 2015 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

AS IV Serial no: 087972  
Version no: 004C

TEST RECORD 00246

Temp Date Time 210L s/

Void: NFI  
12 01/19/14 23:27

Subject Name

Subject I.D.

Operator Name, I.D.

TR. J. L. WHITE 1202  
Location

TROOP D ZONE 3  
ZONE OFFICE  
BOLIVAR, MD

AS IV Serial no: 087972  
Version no: 004C

TEST RECORD 00245

Temp Date Time 210L s/

Air Blank:  
01/19/14 23:26 .000  
Calibration Check:  
21 01/19/14 23:26 .105

Subject Name

Subject I.D.

Operator Name, I.D.

TR. J. L. WHITE 1202  
Location

TROOP D ZONE 3  
ZONE OFFICE  
BOLIVAR, MD

AS IV Serial no: 087972  
Version no: 004C

TEST RECORD 00244

Temp Date Time 210L s/

Air Blank:  
01/19/14 23:25 .000  
Calibration Check:  
20 01/19/14 23:25 .104

Subject Name

Subject I.D.

Operator Name, I.D.

TR. J. L. WHITE 1202  
Location

TROOP D ZONE 3  
ZONE OFFICE  
BOLIVAR, MD

AS IV Serial no: 087972  
Version no: 004C

TEST RECORD 00243

Temp Date Time 210L s/

Air Blank:  
01/19/14 23:23 .000  
Calibration Check:  
20 01/19/14 23:23 .105

Subject Name

Subject I.D.

Operator Name, I.D.

TR. J. L. WHITE #1202  
Location

TROOP D ZONE 3  
ZONE OFFICE  
BOLIVAR, MD