MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

RECEIVED 3/24/14-CD

REVIEWED
By Carol Day at 1:06 pm, May 07, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and when instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
087975	08C.3556.250	03/18/2014
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
119 S. Church St., Union		3:57 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- ☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- ☑ PRINTER WORKING PROPERLY
- ☑ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☑ SIMULATOR SOLUTION ☐ COMPRESSED ETHANOL-GAS MIXTURE

☑ STANDARD SUPPLIER Guth Labs LOT # 14030 EXP. DATE 01/20/2016

☑ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD3332 SIMULATOR EXP DATE 07/10/2014

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☞ .103%	TEST 2 ☞ .103%	TEST 3 ☞ .102%

☑ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
0	0	0	1	2	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Taking instrument out of service to send to factory for software updates

INSPECTING OFFICER

SIGNATURE	PRINT NAME
	Rodney J. Tappe
TYPE & PERMIT NUMBER/EXPIRATION DATE: Expires 09-04-2015	TELEPHONE NUMBER: (636) 583-3700

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2675 James Boulevard
Poplar Bluff, MO 63901

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14030** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **January 22, 2014,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1215%** (w/vol) ethyl alcohol. The expiration date for this lot number is **January 20, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 087975
Version no: 004C

TEST RECORD: 01063

Temp	Date	Time	210L
Blank:			
	03/18/	15:59	.000
Calibration Check:			
	[illegible]	[illegible]	.100

Subject Name

Subject I.D.

Operator Name I.D.

Rod Typ 75

Location

U.P.D

AS IV Serial no: 087975
Version no: 004C

TEST RECORD: 01064

Temp	Date	Time	210L
	[illegible]	[illegible]	[illegible]
Calibration Check:			
	[illegible]	[illegible]	[illegible]

Subject I.D.

Operator Name I.D.

Rod Typ 75

Location

U.P.D

AS IV Serial no: 087975
Version no: 004C

TEST RECORD: 01065

Temp	Date	Time	210L
Blank:			
	[illegible]	16:03	.000
Calibration Check:			
	[illegible]	16:05	[illegible]

Subject I.D.

Operator Name I.D.

Rod Typ 715

Location

U.P.D

AS IV Serial no: 087975
Version no: 004C

TEST RECORD: 01066

Temp	Date	Time	210L
Blank:			
	[illegible]	16:24	

Subject Name

Subject I.D.

Operator Name I.D.

Rod Typ 715

Location

U.P.D

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

RODNEY J TAPPE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/4/2013

NUMBER 230193

EXPIRES 9/4/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Gail Vasterling, acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator TAPPE, RODNEY
Permit No 230193
Date Issued 9/4/2013 Date Expires 9/4/2015