



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

RECEIVED  
 By Carol Day at 8:33 am, Jan 13, 2014

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 087976	PRINTER SN 08C.3527.185	DATE OF INSPECTION 01/09/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 2009 Chevrolet 1/2 ton Truck / 1000 Osage Beach Parkway / Osage Beach, MO		TIME OF INSPECTION 4:38 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION                       COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER RepCo Marketing, Inc.      LOT # 13001      EXP. DATE 03/07/2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 C      SIMULATOR SN SD 3150      SIMULATOR EXP DATE 01/03/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .101%	TEST 2 • .098%	TEST 3 • .102%
----------------	----------------	----------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Calibrate completed.

<b>INSPECTING OFFICER</b>	
SIGNATURE <i>Trooper J.C. Richardson</i>	PRINT NAME Trooper J.C. Richardson
TYPE II PERMIT NUMBER/EXPIRATION DATE 230335 / 12/23/2015	TELEPHONE NUMBER (573) 751-1000

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**  
**LOT NUMBER: 13001**  
**EXPIRATION DATE: March 7, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013  
The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**J C RICHARDSON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, INTOX DMT, ALCO-SENSOR IV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/23/2013

NUMBER 230335

EXPIRES 12/23/2015

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (PS-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator RICHARDSON, J  
Permit No 230335  
Date Issued 12/23/2013 Date Expires 12/23/2015

0\*

AS IV Serial no: 087976  
Version no: 004C

TEST RECORD 00434

Temp Date Time 210L  
Air Blank:  
01/09/14 16:38 .000

Calibration Check:  
26 01/09/14 16:38 .101

Subject Name

Subject I.D.

Operator Name, I.D.

Location

7\*

AS IV Serial no: 087976  
Version no: 004C

TEST RECORD 00435

Temp Date Time 210L  
Air Blank:  
01/09/14 16:40 .000

Calibration Check:  
26 01/09/14 16:40 .098

Subject Name

Subject I.D.

Operator Name, I.D.

Location

0\*

AS IV Serial no: 087976  
Version no: 004C

TEST RECORD 00436

Temp Date Time 210L  
Air Blank:  
01/09/14 16:42 .000

Calibration Check:  
27 01/09/14 16:42 .102

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 087976  
Version no: 004C  
TEST RECORD 00438  
Temp Date Time 210L  
Unit: PFI  
12 01/09/14 16:46  
Subject Name  
Subject I.D.  
Operator Name, I.D.  
Location

AS IV Serial no: 087976  
Version no: 004C  
TEST RECORD 00432  
Temp Date Time 210L  
Air Blank:  
01/09/14 16:32 .000  
Calibration:  
26 01/09/14 16:32 .100  
Subject Name  
Subject I.D.  
Operator Name, I.D.  
Location