



## CERTIFICATE OF ANALYSIS

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**  
**LOT NUMBER: 13001**  
**EXPIRATION DATE: March 7, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013  
The expiration date for this lot number is March 7, 2015 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

AS IV Serial no: 894802  
Version no: 532B

TEST RECORD 00341 s/  
Temp Date Time 210L

Air Blank: 06/20/14 17:34 .000  
Calibration Check: 26 06/20/14 17:34 .102

Subject Name

Subject I.D.

Operator Name, I.D.

DEJMAN 885  
Location

CCPD BOOKING

AS IV Serial no: 894802  
Version no: 532B

TEST RECORD 00342 s/  
Temp Date Time 210L

Air Blank: 06/20/14 17:36 .000  
Calibration Check: 27 06/20/14 17:36 .102

Subject Name

Subject I.D.

Operator Name, I.D.

DEJMAN 885  
Location

CCPD BOOKING

AS IV Serial no: 894802  
Version no: 532B

TEST RECORD 00343 s/  
Temp Date Time 210L

Air Blank: 06/20/14 17:38 .000  
Calibration Check: 27 06/20/14 17:38 .102

Subject Name

Subject I.D.

Operator Name, I.D.

DEJMAN 885  
Location

CCPD BOOKING

AS IV Serial no: 894802  
Version no: 532B

TEST RECORD 00344 s/  
Temp Date Time 210L

VOI: RFI  
12 06/20/14 17:39

Subject Name

Subject I.D.

Operator Name, I.D.

DEJMAN 885  
Location

CCPD BOOKING



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**MICHAEL DEUTMAN**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 5000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 308.111 through 308.119 RSMo.

DATE 12/31/2013  
 NUMBER 230339  
 EXPIRES 12/31/2015

MO 680-0771 (6-10)


*W. S. ...*  
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*David Vashely*  
 acting director  
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 LAB-101

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator DEUTMAN, MICHAEL  
 Permit No 230339  
 Date Issued 12/31/2013 Date Expires 12/31/2015