



CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.  
LOT NUMBER: 14001  
EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014  
The expiration date for this lot number is April 30, 2016 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT  
TYPE II**

**AARON C DEVORSS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, ALCO-SENSOR IV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 06/12/2013  
NUMBER 230115  
EXPIRES 06/12/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Darl Vesterberg*  
Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

MO 580-0771 (6-10)

AS IV Serial no: 094804  
Version no: 532B

TEST RECORD 00202

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:  
09/21/14 08:52 .000  
Calibration Check:  
24 09/21/14 08:52 .098

Subject Name  
Test 1  
Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 094804  
Version no: 532B

TEST RECORD 00204

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:  
09/21/14 08:56 .000  
Calibration Check:  
25 09/21/14 08:56 .098

Subject Name  
Test 3  
Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 094804  
Version no: 532B

TEST RECORD 00206

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:  
09/21/14 09:06 .000  
Subject Test: Auto  
25 09/21/14 09:06 .000

Subject Name  
Practice Test  
Subject I.D.

on myself  
Operator Name, I.D.

Location

AS IV Serial no: 094804  
Version no: 532B

TEST RECORD 00203

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:  
09/21/14 08:54 .000  
Calibration Check:  
25 09/21/14 08:54 .098

Subject Name  
Test 2  
Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 094804  
Version no: 532B

TEST RECORD 00205

Temp	Date	Time	s/	210L
------	------	------	----	------

VOID: RFI  
12 09/21/14 08:57

Subject Name  
RFI Test  
Subject I.D.

Operator Name, I.D.

Location