

AS IV Serial no: 094804
Version no: 532B

TEST RECORD 00207

Temp Date Time ^{s/} 210L

Air Blank:
10/23/14 12:04 .000
Calibration Check:
22 10/23/14 12:04 .100

Subject Name

TEST 1

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 094804
Version no: 532B

TEST RECORD 00209

Temp Date Time ^{s/} 210L

Air Blank:
10/23/14 12:08 .000
Calibration Check:
24 10/23/14 12:08 .099

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 094804
Version no: 532B

TEST RECORD 00208

Temp Date Time ^{s/} 210L

Air Blank:
10/23/14 12:06 .000
Calibration Check:
23 10/23/14 12:06 .101

Subject Name

TEST 2

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 094804
Version no: 532B

TEST RECORD 00211

Temp Date Time ^{s/} 210L

VOID: RFI
12 10/23/14 12:12

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

Location

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 14001
EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014
The expiration date for this lot number is April 30, 2016 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

AARON C DEVORSS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 06/12/2013

NUMBER 230115

EXPIRES 06/12/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Wm W. Devorss
 Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES