



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

REPORT #7

By Carol Day at 8:02 am, Jan 02, 2015

Complete this report in duplicate at the time of the regular monthly preventative maintenance check and whenever instrument is repaired.
Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 094804	PRINTER SN 097.3584.334	DATE OF INSPECTION 12/31/2014
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 201 W. 2nd St. Sedalia, MO 65301	TIME OF INSPECTION 5:24 pm
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER REPCO LOT # 14001 EXP. DATE 04/30/2016

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2306 SIMULATOR EXP DATE 10/23/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 0.097	TEST 2 0.097	TEST 3 0.097
---------------	---------------	---------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

MEMORY PRINTED. MEMORY ERASED. DEVICE TAKEN OUT OF SERVICE.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME AARON CASEY DEVORSS
---------------	-----------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 230115 06.12.2015	TELEPHONE NUMBER (660) 826-8100
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 14001
EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014
The expiration date for this lot number is April 30, 2016 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

AARON C DEVORSS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 06/12/2013

NUMBER 230115

EXPIRES 06/12/2015

W. S. R.

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paul Vesterberg
Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

AS IV Serial no: 094004
Version no: 532B

Last Calibration:
06/23/14 11:36 .100

Test Results:

00169 Calibration:
25 04/23/14 13:45 .100
00.0 Lit. 00.0 Sec.
00170 Calibration Check:
26 04/23/14 13:49 .101
00.0 Lit. 00.0 Sec.
00171 Calibration Check:
26 04/23/14 13:50 .101
00.0 Lit. 00.0 Sec.
00172 Calibration Check:
27 04/23/14 13:52 .100
00.0 Lit. 00.0 Sec.
00173 VOID: RFI
12 04/23/14 13:53
00.0 Lit. 00.0 Sec.
00174 Subject Test: Man
22 05/23/14 11:32 .000
00.0 Lit. 00.0 Sec.
00175 Calibration Check:
23 05/23/14 11:34 .105
00.0 Lit. 00.0 Sec.
00176 Calibration Check:
25 05/23/14 11:37 .103
00.0 Lit. 00.0 Sec.
00177 Calibration Check:
25 05/23/14 11:39 .103
00.0 Lit. 00.0 Sec.
00178 Calibration Check:
26 05/23/14 11:41 .103
00.0 Lit. 00.0 Sec.
00179 VOID: RFI
12 05/23/14 11:43
00.0 Lit. 00.0 Sec.
00180 Subject Test: Auto
22 06/06/14 10:05 .000
02.9 Lit. 10.9 Sec.
00181 Subject Test: Auto
22 06/06/14 11:04 .000
01.6 Lit. 07.5 Sec.
00182 Subject Test: Auto
23 06/06/14 11:06 .000
02.0 Lit. 06.0 Sec.
00183 Calibration Check:
22 06/23/14 11:30 .102
00.0 Lit. 00.0 Sec.
00184 Calibration Check:
23 06/23/14 11:32 .105
00.0 Lit. 00.0 Sec.
00185 Calibration Check:

24 06/23/14 11:34 .107
00.0 Lit. 00.0 Sec.
00186 Calibration:
25 06/23/14 11:36 .100
00.0 Lit. 00.0 Sec.
00187 Calibration Check:
26 06/23/14 11:39 .101
00.0 Lit. 00.0 Sec.
00188 Calibration Check:
26 06/23/14 11:40 .100
00.0 Lit. 00.0 Sec.
00189 Calibration Check:
27 06/23/14 11:42 .101
00.0 Lit. 00.0 Sec.
00190 VOID: RFI
12 06/23/14 11:43
00.0 Lit. 00.0 Sec.
00191 Subject Test: Man
22 07/23/14 09:38 .000
00.0 Lit. 00.0 Sec.
00192 Calibration Check:
23 07/23/14 09:40 .100
00.0 Lit. 00.0 Sec.
00193 Calibration Check:
24 07/23/14 09:42 .100
00.0 Lit. 00.0 Sec.
00194 Calibration Check:
25 07/23/14 09:43 .099
00.0 Lit. 00.0 Sec.
00195 VOID: RFI
12 07/23/14 09:44
00.0 Lit. 00.0 Sec.
00196 Subject Test: Man
23 08/21/14 09:44 .000
00.0 Lit. 00.0 Sec.
00197 Calibration Check:
24 08/21/14 09:46 .100
00.0 Lit. 00.0 Sec.
00198 Calibration Check:
25 08/21/14 09:47 .099
00.0 Lit. 00.0 Sec.
00199 Calibration Check:
26 08/21/14 09:50 .100
00.0 Lit. 00.0 Sec.
00200 VOID: RFI
12 08/21/14 09:51
00.0 Lit. 00.0 Sec.
00201 VOID: SET PUSH
8 09/21/14 08:51
00.0 Lit. 00.0 Sec.
00202 Calibration Check:

24 09/21/14 08:52 .098
00.0 Lit. 00.0 Sec.
00203 Calibration Check:
25 09/21/14 08:54 .098
00.0 Lit. 00.0 Sec.
00204 Calibration Check:
25 09/21/14 08:56 .098
00.0 Lit. 00.0 Sec.
00205 VOID: RFI
12 09/21/14 08:57
00.0 Lit. 00.0 Sec.
00206 Subject Test: Auto
25 09/21/14 09:06 .000
01.5 Lit. 04.6 Sec.
00207 Calibration Check:
22 10/23/14 12:04 .100
00.0 Lit. 00.0 Sec.
00208 Calibration Check:
23 10/23/14 12:06 .101
00.0 Lit. 00.0 Sec.
00209 Calibration Check:
24 10/23/14 12:08 .099
00.0 Lit. 00.0 Sec.
00210 Subject Test: Man
24 10/23/14 12:10 .074
00.0 Lit. 00.0 Sec.
00211 VOID: RFI
12 10/23/14 12:12
00.0 Lit. 00.0 Sec.
00212 Calibration Check:
19 11/29/14 01:11 .000
00.0 Lit. 00.0 Sec.
00213 Calibration Check:
20 11/29/14 01:15 .100
00.0 Lit. 00.0 Sec.
00214 Calibration Check:
21 11/29/14 01:16 .100
00.0 Lit. 00.0 Sec.
00215 Calibration Check:
22 11/29/14 01:18 .105
00.0 Lit. 00.0 Sec.
00216 Calibration Check:
23 11/29/14 01:19 .101
00.0 Lit. 00.0 Sec.
00217 VOID: RFI
12 11/29/14 01:21
00.0 Lit. 00.0 Sec.

MEMORY PRINT

AS IV Serial no: 094804
Version no: 532B

TEST RECORD 00218

Temp Date Time ^{s/} 210L

Air Blank:
12/31/14 17:24 .000
Calibration Check:
22 12/31/14 17:24 .097

Subject Name

Test 1
Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 094804
Version no: 532B

TEST RECORD 00219

Temp Date Time ^{s/} 210L

Air Blank:
12/31/14 17:26 .000
Calibration Check:
23 12/31/14 17:26 .097

Subject Name

Test 2
Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 094804
Version no: 532B

TEST RECORD 00220

Temp Date Time ^{s/} 210L

Air Blank:
12/31/14 17:28 .000
Calibration Check:
23 12/31/14 17:28 .097

Subject Name

Test 3
Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 094804
Version no: 532B

TEST RECORD 00221

Temp Date Time ^{s/} 210L

VOID: RFI
12 12/31/14 17:30

Subject Name

RFI
Subject I.D.

Operator Name, I.D.

Location