

AS IV Serial no: 095961
Version no: 532B

TEST RECORD 00338

Temp	Date	Time	a/ 210L
------	------	------	------------

Air Blank:	04/22/14	08:19	.000
Calibration Check:	21	04/22/14	08:19 .102

Subject Name

N/A

Subject I.D.

N/A

Operator Name, I.D.

E. DYE

Location

715 W. Center Circle

Nixa Mo 65714

AS IV Serial no: 095961
Version no: 532B

TEST RECORD 00339

Temp	Date	Time	a/ 210L
------	------	------	------------

Air Blank:	04/22/14	08:21	.000
Calibration Check:	22	04/22/14	08:21 .102

Subject Name

N/A

Subject I.D.

N/A

Operator Name, I.D.

E. DYE 415

Location

715 W. Center Circle

Nixa Mo 65714

AS IV Serial no: 095961
Version no: 532B

TEST RECORD 00337

Temp	Date	Time	a/ 210L
------	------	------	------------

Air Blank:	04/22/14	08:17	.000
Calibration Check:	20	04/22/14	08:17 .102

Subject Name

N/A

Subject I.D.

N/A

Operator Name, I.D.

E. DYE 415

Location

715 W Center Circle

Nixa Mo 65714

AS IV Serial no: 095961
Version no: 532B

TEST RECORD 00340

Temp	Date	Time	a/ 210L
------	------	------	------------

VOID: RFI	12	04/22/14	08:23
-----------	----	----------	-------

Subject Name

RFI TEST

Subject I.D.

N/A

Operator Name, I.D.

E. DYE 415

Location

715 W Center Circle

Nixa Mo 65714

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II
ERIC DYE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/17/2013

NUMBER 230316

EXPIRES 12/17/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator DYE, ERIC
Permit No 230316
Date Issued 12/17/2013 Date Expires 12/17/2015