



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 3:58 pm, Nov 10, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097413	PRINTER SN 091.3580.067	DATE OF INSPECTION 11/04/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 52 YOUNG DR CALVERTON PARK MO 63135		TIME OF INSPECTION 1:19 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH LABORATORIES LOT # 14030 EXP. DATE 01/20/2016

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN SD3327 SIMULATOR EXP DATE 01/13/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 \Rightarrow .099

TEST 2 \Rightarrow .099

TEST 3 \Rightarrow .098

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

GUTH LAB .100 EXP 01/20/2016 LOT NUMBER 14030 TIME CHANGE

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Lowell Moore
TYPE II PERMIT NUMBER/EXPIRATION DATE 240234 05/12/2016	TELEPHONE NUMBER (314) 524-1212

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14030 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 22, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1215% (w/vol) ethyl alcohol. The expiration date for this lot number is January 20, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 097413
Version no: 004C

TEST RECORD 00496 s/
Temp Date Time 210L

Air Blank: 11/04/14 13:19 .000
Calibration Check: 22 11/04/14 13:19 .099

Subject Name

Subject I.D.

L. Moore 290234
Operator Name, I.D.

Location

d

AS IV Serial no: 097413
Version no: 004C

TEST RECORD 00497 s/
Temp Date Time 210L

Air Blank: 11/04/14 13:21 .000
Calibration Check: 22 11/04/14 13:21 .099

Subject Name

Subject I.D.

L. Moore 290234
Operator Name, I.D.

Location

3

AS IV Serial no: 097413
Version no: 004C

TEST RECORD 00498 s/
Temp Date Time 210L

Air Blank: 11/04/14 13:22 .000
Calibration Check: 23 11/04/14 13:22 .098

Subject Name

Subject I.D.

L. Moore 290239
Operator Name, I.D.

Location

RFI

AS IV Serial no: 097413
Version no: 004C

TEST RECORD 00499

Temp Date Time 210L
9/

Void: RFI
12 11/04/14 13:24

Subject Name

Subject I.D.

Jm 240239

Operator Name, I.D.

Location



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
 TYPE II

LOWELL J MOORE

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/12/2014

NUMBER 240234

EXPIRES 5/12/2016

[Signature]
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

[Signature]
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES


LAB-4 (8-10)

MO 580-0771 (6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator MOORE, LOWELL
 Permit No 240234
 Date Issued 5/12/2014 Date Expires 5/12/2016