





**GUTH LABORATORIES, INC.**

550 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13290 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1202% (w/vol) ethyl alcohol. The expiration date for this lot number is October 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

AS IV Serial no: 097416  
Version no: 004C

TEST RECORD - REPRINT

TEST RECORD 00206

Temp Date Time 210L

AS IV Serial no: 097416  
Version no: 004C

TEST RECORD - REPRINT

TEST RECORD 00206

Temp Date Time 210L

Air Blank:  
05/26/14 11:23 .000  
Calibration Check:  
25 05/26/14 11:23 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 097416  
Version no: 004C

TEST RECORD 00204

Temp Date Time 210L

Air Blank:  
05/26/14 11:19 .000  
Calibration Check:  
24 05/26/14 11:19 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 097416  
Version no: 004C

TEST RECORD 00205

Temp Date Time 210L

Air Blank:  
05/26/14 11:21 .000  
Calibration Check:  
25 05/26/14 11:21 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 097416  
Version no: 004C

TEST RECORD 00207

Temp Date Time 210L

Void: RFI  
12 05/26/14 11:49

Subject Name

Subject I.D.

Operator Name, I.D.

Location



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT  
TYPE II**

**JOSEY J LONG**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, INTOX DMT, ALCO-SENSOR IV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014

NUMBER 240062

EXPIRES 3/7/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0171 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator **LONG, JOSEY**  
Permit No **240062**  
Date Issued **3/7/2014**    Date Expires **3/7/2016**