



AS IV Serial no: 097419  
Version no: 532B

TEST RECORD 00062

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
08/01/14 06:19 .000  
Calibration Check:  
21 08/01/14 06:19 .100

Subject Name  
CALIBRATION CHECK  
Subject I.D.

Operator Name, I.D.  
Ronald Chery 501  
Location:  
Palmira PD

AS IV Serial no: 097419  
Version no: 532B

TEST RECORD 00063

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
08/01/14 06:21 .000  
Calibration Check:  
21 08/01/14 06:21 .100

Subject Name  
CALIBRATION CHECK  
Subject I.D.

Operator Name, I.D.  
Ronald Chery 501  
Location:  
Palmira PD

AS IV Serial no: 097419  
Version no: 532B

TEST RECORD 00064

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
08/01/14 06:22 .000  
Calibration Check:  
22 08/01/14 06:22 .101

Subject Name  
CALIBRATION CHECK  
Subject I.D.

Operator Name, I.D.  
Ronald Chery 501  
Location:  
Palmira PD

AS IV Serial no: 097419  
Version no: 532B

TEST RECORD 00065

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 08/01/14 06:23

Subject Name  
RFI TEST  
Subject I.D.

Operator Name, I.D.  
Ronald Chery 501  
Location:  
Palmira PD



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14110 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 5, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is May 1, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**RONALD C PEER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/22/2014

NUMBER 240306

EXPIRES 7/22/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO:580-0771 (6-10)

LAB-4 (RS-10)



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator PEER, RONALD  
 Permit No 240306  
 Date Issued 7/22/2014 Date Expires 7/22/2016



## Simulator Calibration Report

This calibration report is to certify the alcohol reference simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance with the standards set by the Missouri Department of Health and Senior Services Rules and Regulations:  
19 CSR 25-30.051 (4).



Technician Printed Name: DONALD D. DEBUONO

Technician Signature: Donald D. DeBuono

Date: 07/14/2014

Contact: Missouri Safety Center

Breath-Alcohol Instrument Training Program

660-543-4834