



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever the instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

RECEIVED
 By Carol Day at 4:04 pm, Jan 15, 2014

ALCO SENSOR IV SN 097421	PRINTER SN 096.3580.987	DATE OF INSPECTION 01/13/2014
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 208 E. Pearl Harrisonville	TIME OF INSPECTION 12:44 pm
--	--------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Repco Marketing LOT # 13002 EXP. DATE 06/19/2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2254 SIMULATOR EXP DATE 01/13/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .105

TEST 2 ← .104

TEST 3 ← .105

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	1	(0-.04)	1	(.05-.09)	0	(.10-.14)	3	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Michael Davis
---------------	-----------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 230250 10/31/2015	TELEPHONE NUMBER (816) 380-8940
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

Operator Name, I.D.

Subject I.D.

72513

Subject Name

01/13/14 12:47.000
Air Blank
Calibration Check: 18 01/13/14 12:47.185

TEST RECORD 00554
9/
Temp Date Time 210L

NS IV Serial no: 092421
Version not 004C

Location

Operator Name, I.D.

Subject I.D.

7F2 TEST

Subject Name

01/13/14 12:49.000
Air Blank
Calibration Check: 12 01/13/14 12:49.185

TEST RECORD 00555
9/
Temp Date Time 210L

NS IV Serial no: 092421
Version not 004C

Location

Operator Name, I.D.

Subject I.D.

TEST 2

Subject Name

01/13/14 12:46.000
Air Blank
Calibration Check: 18 01/13/14 12:46.184

TEST RECORD 00553
9/
Temp Date Time 210L

NS IV Serial no: 092421
Version not 004C

Location

Operator Name, I.D.

Subject I.D.

TEST 1

Subject Name

01/13/14 12:44.000
Air Blank
Calibration Check: 17 01/13/14 12:44.185

TEST RECORD 00552
9/
Temp Date Time 210L

NS IV Serial no: 092421
Version not 004C

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

MICHAEL L DAVIS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 5000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/31/2013

NUMBER 230250

EXPIRES 10/31/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Gal Vasterby, acting director
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (16-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator DAVIS, MICHAEL
 Permit No 230250
 Date Issued 10/31/2013 Date Expires 10/31/2015