



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 12:14 pm, Oct 28, 2014

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097421	PRINTER SN 096.3580.987	DATE OF INSPECTION 10/23/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 110 Silverlady Ln Branson West Mo, 65737		TIME OF INSPECTION 3:55 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth Laboratories LOT # 14030 EXP. DATE 01/20/2016
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD3139 SIMULATOR EXP DATE 07/30/2015

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .101	TEST 2 ➔ .103	TEST 3 ➔ .103
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Vincent Collins
TYPE II PERMIT NUMBER/EXPIRATION DATE 240349 10-03-2016	TELEPHONE NUMBER (417) 272-3400

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

V


CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14030 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 22, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1215% (w/vol) ethyl alcohol. The expiration date for this lot number is January 20, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.



Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
VINCENT COLLINS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/3/2014

NUMBER 240349

EXPIRES 10/3/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator COLLINS, VINCENT
 Permit No 240349
 Date Issued 10/3/2014 Date Expires 10/3/2016

AS IV Serial no: 897421
Version no: 532B

TEST RECORD 00080

Temp Date Time ^{s/} 210L

Air Blank:
16/23/14 15:55 .000
Calibration Check:
26 16/23/14 15:55 .101

Subject Name

Test #1

Subject I.D.

Vincent Collins

Operator Name, I.D.

240349

Location

Branson West Police Dept

110 Silverlady Lane
Branson West mo 65737

VERSION NO: 532B

TEST RECORD 00081

Temp Date Time ^{s/} 210L

Air Blank:
16/23/14 15:56 .000
Calibration Check:
26 16/23/14 15:56 .103

Subject Name

Test #2

Subject I.D.

Vincent Collins

Operator Name, I.D.

240349

Location

Branson West Police Dept.

110 Silverlady Ln
Branson West mo
65737

AS IV Serial no: 897421
Version no: 532B

TEST RECORD 00082

Temp Date Time ^{s/} 210L

Air Blank:
16/23/14 15:58 .000
Calibration Check:
27 16/23/14 15:58 .103

Subject Name

Test #3

Subject I.D.

Vincent Collins

Operator Name, I.D.

240349

Location

110 Silverlady Ln

Branson West mo 65737

AS IV Serial no: 897421
Version no: 532B

TEST RECORD 00083

Temp Date Time ^{s/} 210L

VOID: RFI
12 16/23/14 15:59

Subject Name

RFI Test

Subject I.D.

Vincent Collins

Operator Name, I.D.

240349

Location

110 Silverlady Ln

Branson West mo 65737

AS IV Serial no: 897421
Version no: 532B

TEST RECORD 00084

Temp Date Time ^{s/} 210L

Air Blank:
16/23/14 16:01 .000
Subject Test: Auto
28 16/23/14 16:01 .000

Subject Name

Blank Test

Subject I.D.

Vincent Collins

Operator Name, I.D.

240349

Location

110 Silverlady Ln

Branson West mo