



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-554-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14200 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on August 6, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1213% (w/vol) ethyl alcohol. The expiration date for this lot number is August 5, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS 1U Serial no: 097421
Version no: 532B

TEST RECORD 00098

Temp Date Time 210L

VOID: RFI
12 12/06/14 18:05

Subject Name

RFI / Test

Subject I.D.

Vincent Collins 303

Operator Name, I.D.

240349 10-3-16

Location

110 Silverlady Ln

Branson West mo. 65737

Vincent Collins

AS 1U Serial no: 097421
Version no: 532B

TEST RECORD 00099

Temp Date Time 210L

Air Blank:
12/06/14 18:06 .000
Subject Test: Auto.
24 12/06/14 18:06 .000

Subject Name

Blank / Test

Subject I.D.

Vincent Collins 303

Operator Name, I.D.

240349 10-3-16

Location

110 Silverlady Ln

Branson West mo. 65737

Vincent Collins

AS IU Serial no: 097421
Version no: 532B

AS IU Serial no: 097421
Version no: 532B

TEST RECORD 00095

TEST RECORD 00095
Temp Date Time 210L
Air Blank:
12/06/14 18:01 .000
Calibration Check:
21 12/06/14 18:01 .102

TEST RECORD 00096
Temp Date Time 210L
Air Blank:
12/06/14 18:02 .000
Calibration Check:
23 12/06/14 18:02 .100

Subject Name
Test 1
Subject I.D.
Vincent Collins 303
Operator Name, I.D.
240349 10-3-16
Location
110 Silverlady Ln

Subject Name
Test 2
Subject I.D.
Vincent Collins 303
Operator Name, I.D.
240349 10-3-16
Location
110 Silverlady Ln

Branson West mo 65737

David Callio

Branson West mo 65737

David Callio

AS IU Serial no: 097421
Version no: 532B

TEST RECORD 00097

Temp Date Time 210L
Air Blank:
12/06/14 18:04 .000
Calibration Check:
23 12/06/14 18:04 .101

Subject Name
Test 3
Subject I.D.
Vincent Collins 303
Operator Name, I.D.
240349 10-3-16
Location
110 Silverlady Ln

Branson West mo 65737

David Callio



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

VINCENT COLLINS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, DATAMASTER


for the determination of the alcoholic content of blood from a sample of expired air, Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/3/2014

NUMBER 240349

EXPIRES 10/3/2016


DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


, acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680-0771 (8-10)

LAB-4 (15-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator COLLINS, VINCENT
Permit No 240349
Date issued 10/3/2014 Date Expires 10/3/2016