



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097425	PRINTER SN 096.3580.866	DATE OF INSPECTION 07/03/2014 8/4/14-cd
LOCATION OF INSTRUMENT (STREET AND CITY) Troop F Headquarters		TIME OF INSPECTION 8:33 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth Laboratories, Inc. LOT # 13290 EXP. DATE 10/29/2015
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN CD05457 SIMULATOR EXP DATE 06/03/2015
- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .098	TEST 2 .098	TEST 3 .099
--------------	--------------	--------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Trooper R.R. Dowd
TYPE II PERMIT NUMBER/EXPIRATION DATE 240272 / 06-13-2016	TELEPHONE NUMBER (573) 751-1000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13290** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 31, 2013**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1202%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 29, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** $\pm 3\%$.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 097425
Version no: 004C

TEST RECORD 00248

Temp Date Time ^{s/} 210L

Air Blank:
08/04/14 20:33 .000

Calibration Check:
21 08/04/14 20:33 .098

Subject Name

TEST # 1

Subject I.D.

Operator Name, I.D.

TPR RR Dowd

Location

TROOP F HEADQUARTERS

AS IV Serial no: 097425
Version no: 004C

TEST RECORD 00249

Temp Date Time ^{s/} 210L

Air Blank:
08/04/14 20:36 .000

Calibration Check:
22 08/04/14 20:36 .098

Subject Name

TEST # 2

Subject I.D.

Operator Name, I.D.

TPR RR Dowd

Location

TROOP F HEADQUARTERS

AS IV Serial no: 097425
Version no: 004C

TEST RECORD 00250

Temp Date Time ^{s/} 210L

Air Blank:
08/04/14 20:38 .000

Calibration Check:
23 08/04/14 20:38 .099

Subject Name

TEST # 3

Subject I.D.

Operator Name, I.D.

TPR RR Dowd

Location

TROOP F HEADQUARTERS

AS IV Serial no: 097425
Version no: 004C

TEST RECORD 00252

Temp Date Time ^{s/} 210L

Void: RFI
12 08/04/14 20:56

Subject Name

RFI TEST

Subject I.D.

Operator Name, I.D.

TPR RR Dowd

Location

TROOP F HEADQUARTERS