



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

By Carol Day at 9:09 am, Sep 08, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097426	PRINTER SN 03A. 2436. 098	DATE OF INSPECTION 08/25/14
LOCATION OF INSTRUMENT (STREET AND CITY) 3310 NE RENNAN DR., LEE'S SUMMIT, MO - 64064		TIME OF INSPECTION 19:50

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C) **22° C**

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER **REPCO MARKETING** LOT # **13002** EXP. DATE **06/19/2015**

SIMULATOR TEMPERATURE (34°C ± 0.2°C) **34.0°C** SIMULATOR SN **SD3329** SIMULATOR EXP DATE **10/03/2014**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 = .104	TEST 2 = .105	TEST 3 = .104
----------------------	----------------------	----------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	1	(.10-.14)	2	(.15-.19)	1	(OVER .19)	0
----------	----------	---------	----------	-----------	----------	-----------	----------	-----------	----------	------------	----------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSTRUMENT MEETS ALL DOHSS STANDARDS AND GUIDELINES

INSPECTING OFFICER

SIGNATURE 	PRINT NAME T. HERRMANN
TYPE II PERMIT NUMBER/EXPIRATION DATE # 230202 09/19/2015	TELEPHONE NUMBER (816) - 524 - 4302

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 097426
Version no: 532B

TEST RECORD 00086
%
Temp Date Time 210L

Air Blank:
08/25/14 19:50 .000
Calibration Check:
22 08/25/14 19:50 .104

Subject Name
MONTHLY CERT

Subject I.D.
HERRMANN #77

Operator Name, I.D.
#230202 09/19/15

Location
3310 NE RENNAU DR

CEE'S SUMMIT, MO

AS IV Serial no: 097426
Version no: 532B

TEST RECORD 00087
%
Temp Date Time 210L

Air Blank:
08/25/14 19:52 .000
Calibration Check:
22 08/25/14 19:52 .105

Subject Name
MONTHLY CERT

Subject I.D.
HERRMANN #77

Operator Name, I.D.
#230202 09/19/15

Location
3310 NE RENNAU DR

CEE'S SUMMIT, MO

AS IV Serial no: 097426
Version no: 532B

TEST RECORD 00088
%
Temp Date Time 210L

Air Blank:
08/25/14 19:54 .000
Calibration Check:
23 08/25/14 19:54 .104

Subject Name
MONTHLY CERT

Subject I.D.
HERRMANN #77

Operator Name, I.D.
#230202 09/19/15

Location
3310 NE RENNAU DR

CEE'S SUMMIT, MO

AS IV Serial no: 097426
Version no: 532B

TEST RECORD 00089
%
Temp Date Time 210L

VOID: RFI
12 08/25/14 19:55

Subject Name
MONTHLY CERT

Subject I.D.
HERRMANN #77

Operator Name, I.D.
#230202 09/19/15

Location
3310 NE RENNAU DR

CEE'S SUMMIT, MO

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

TRAVIS HERRMANN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/19/2013

NUMBER 230202

EXPIRES 9/19/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator HERRMANN, TRAVIS
 Permit No 230202
 Date Issued 9/19/2013 Date Expires 9/19/2015