

AS IU Serial no: 097429
Version no: 532B

TEST RECORD 00012

Temp Date Time 210L
9/

Air Blank:

05/25/14 06:43 .000

Calibration Check:

20 05/25/14 06:43 .095

Subject Name

Weatherly, Marc

Subject I.D.

5/25/14

Operator Name, I.D.

Roger Wastler 240140

Location

34 Dillow Plaza

High Ridge

AS IU Serial no: 097429
Version no: 532B

TEST RECORD 00011

Temp Date Time 210L
9/

Air Blank:

05/25/14 06:41 .000

Calibration Check:

20 05/25/14 06:41 .095

Subject Name

Weatherly, Marc

Subject I.D.

5/25/14

Operator Name, I.D.

Roger Wastler 240140

Location

34 Dillow Plaza

High Ridge

AS IU Serial no: 097429
Version no: 532B

TEST RECORD 00010

Temp Date Time 210L
9/

Air Blank:

05/25/14 06:40 .000

Calibration Check:

19 05/25/14 06:40 .095

Subject Name

Weatherly, Marc

Subject I.D.

5/25/14

Operator Name, I.D.

Roger Wastler 240140

Location

34 Dillow Plaza

High Ridge

AS IU Serial no: 097429
Version no: 532B

TEST RECORD 00013

Temp Date Time 210L
9/

VOID: RFI

12 05/25/14 06:44

Subject Name

RFI Test

Subject I.D.

5/25/14

Operator Name, I.D.

Roger Wastler 240140

Location

34 Dillow Plaza

High Ridge



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14110 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 5, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is May 1, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.


The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.


NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

ROGER WENSLER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX EC/IR II, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/3/2014

NUMBER 240140

EXPIRES 4/3/2016

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paul Vonderly

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)