



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

**RECEIVED**  
 By Carol Day at 11:42 am, Dec 08, 2014 red.

Complete this report in duplicate at the time of the regular monthly preventative maintenance check. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097439	PRINTER SN 096.3580.990	DATE OF INSPECTION 12/05/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 2014 Ford Interceptor, Osage Beach Parkway, Osage Beach		TIME OF INSPECTION 11:47 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories, Inc. LOT # 13280 EXP. DATE 10/16/2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN SD 3150 SIMULATOR EXP DATE 01/03/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .105%

TEST 2 ➡ .102%

TEST 3 ➡ .105%

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE *Trooper C. Richardson* (253)

PRINT NAME  
Trooper C. Richardson

TYPE II PERMIT NUMBER/EXPIRATION DATE  
230335 12/23/2015

TELEPHONE NUMBER  
(573) 751-1000

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901



**GUTH LABORATORIES, INC.**

800 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

AS IV Serial no: 097439  
Version no: 004C

TEST RECORD 00199

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
12/05/14 11:47 .000  
Calibration Check:  
16 12/05/14 11:47 .105

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 097439  
Version no: 004C

TEST RECORD 00201

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
12/05/14 11:50 .000  
Calibration Check:  
19 12/05/14 11:50 .102

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 097439  
Version no: 004C

TEST RECORD 00202

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
12/05/14 11:52 .000  
Calibration Check:  
20 12/05/14 11:52 .105

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 097439  
Version no: 004C

TEST RECORD 00203

Temp Date Time 210L <sup>s/</sup>

Void: RFI  
12 12/05/14 11:53

Subject Name

Subject I.D.

Operator Name, I.D.

Location