MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 4:25 pm, Apr 23, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
097444	096.3580.952	04/19/2014
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
9623 SAINT CHARLES ROCK ROAD, BRECKENRIDGE HILLS POLICE DEPARTMENT		5:49 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- ☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C) 26°C
- ☑ PRINTER WORKING PROPERLY
- ☑ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☑ SIMULATOR SOLUTION ☐ COMPRESSED ETHANOL-GAS MIXTURE

☑ STANDARD SUPPLIER GUTH LABORATORIES, INC. LOT # 14030 EXP. DATE 01/20/2016

☑ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2309 SIMULATOR EXP DATE 12/16/2014

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛ .103	TEST 2 ☛ .103	TEST 3 ☛ .103

☑ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE	PRINT NAME
Sgt Michael Presson 233	MICHAEL PRESSON
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
230108/06-04-2015	(314) 426-1214

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14030** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **January 22, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1215%** (w/vol) ethyl alcohol. The expiration date for this lot number is **January 20, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 097444
Version no: 004C

TEST RECORD 00863

Temp	Date	Time	9/ 210L
Air Blank:			
	04/19/14	17:49	.000
Subject Test Man			
25	04/19/14	17:49	.103

Subject Name

Subject I.D.

Operator Name, I.D.
SGT Michael Pierson 232

Location
9623 St. Charles Rock Rd

AS IV Serial no: 097444
Version no: 004C

TEST RECORD 00864

Temp	Date	Time	9/ 210L
Air Blank:			
	04/19/14	17:50	.000
Subject Test Man			
26	04/19/14	17:50	.103

Subject Name

Subject I.D.

Operator Name, I.D.
SGT Michael Pierson 232

Location
9623 St. Charles Rock Rd

AS IV Serial no: 097444
Version no: 004C

TEST RECORD 00865

Temp	Date	Time	9/ 210L
Air Blank:			
	04/19/14	17:51	.000
Subject Test Man			
27	04/19/14	17:51	.103

Subject Name

Subject I.D.

Operator Name, I.D.
SGT Michael Pierson 232

Location
9623 St. Charles Rock Rd

AS IV Serial no: 097444
Version no: 004C

TEST RECORD 00866

Temp	Date	Time	9/ 210L
Weight Est.			
12	04/19/14	17:50	

Subject Name

Subject I.D.

Operator Name, I.D.
SGT Michael Pierson 232

Location
9623 St. Charles Rock Rd

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

MICHAEL B PRESSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 06/04/2013

NUMBER 230108

EXPIRES 06/04/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Gail Vasterling

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)