


 MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 10:43 am, Apr 14, 2014

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097445	PRINTER SN 096.3580.951	DATE OF INSPECTION 04/11/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 300 NORTH NEW BALLAS ROAD, CREVE COEUR		TIME OF INSPECTION 5:51 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

 DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

 TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

 PRINTER WORKING PROPERLY

 TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

 SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

 STANDARD SUPPLIER REPCO MARKETING INC. LOT # 13001 EXP. DATE 03/07/2015
 SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN SD2784 SIMULATOR EXP DATE 07/10/2014
 CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .099%TEST 2 .100%TEST 3 .100%
 RFI DETECTOR OPERATING

 INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 | (0-.04) 0 | (.05-.09) 0 | (.10-.14) 0 | (.15-.19) 2 | (OVER .19) 1

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE <i>P.O. Michael J. Deutman #885</i>	PRINT NAME Michael Deutman #885
TYPE II PERMIT NUMBER/EXPIRATION DATE 230339 12/31/2015	TELEPHONE NUMBER (314) 872-0946

 Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

RepCo Marketing Inc.

3101-189 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-878-3480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13001
EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013.
The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

AS IV Serial no: 097445
 Version no: 532B

TEST RECORD 06022

Temp	Date	Time	%
Air Blank:	04/11/14	17:53	.286
Calibration Check:	19 04/11/14	17:53	.166

Subject Name

Subject I.D.

Operator Name, I.D.
 DEJMAN 885

Location
 CCPD BATVAN

AS IV Serial no: 097445
 Version no: 532B

TEST RECORD 06023

Temp	Date	Time	%
Air Blank:	04/11/14	17:55	.288
Calibration Check:	20 04/11/14	17:55	.168

Subject Name

Subject I.D.

Operator Name, I.D.
 DEJMAN 885

Location
 CCPD BATVAN

AS IV Serial no: 097445
 Version no: 532B

TEST RECORD 06024

Temp	Date	Time	%
Air Blank:	04/11/14	17:05	

Subject Name

Subject I.D.

Operator Name, I.D.
 DEJMAN 885

Location
 CCPD BATVAN

AS IV Serial no: 097445
 Version no: 532B

TEST RECORD 06021

Temp	Date	Time	%
Air Blank:	04/11/14	17:51	.286
Calibration Check:	19 04/11/14	17:51	.099

Subject Name

Subject I.D.

Operator Name, I.D.
 DEJMAN 885

Location
 CCPD BATVAN



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

MICHAEL DEUTMAN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 5000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 308.111 through 308.119 RSMo.

DATE 12/31/2013

NUMBER 230339

EXPIRES 12/31/2015

MO 680-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (98-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator DEUTMAN, MICHAEL
Permit No 230339
Date Issued 12/31/2013 Date Expires 12/31/2015