

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

AS IV Serial no: 897445
Version no: 5028

TEST RECORD 80836 s/
Temp Date Time 210L
Air Blank: 06/28/14 17:26 .880
Calibration Check: 19 06/28/14 17:26 .897

Subject Name

Subject I.D.

Operator Name, I.D.
DEUTMAN 883

Location
CCPD BATVAN

AS IV Serial no: 897445
Version no: 5028

TEST RECORD 80839 s/
Temp Date Time 210L
Air Blank: 06/28/14 17:28 .880
Calibration Check: 20 06/28/14 17:28 .896

Subject Name

Subject I.D.

Operator Name, I.D.
DEUTMAN 883

Location
CCPD BATVAN

AS IV Serial no: 897445
Version no: 5028

TEST RECORD 80840 s/
Temp Date Time 210L
Air Blank: 06/28/14 17:28 .880
Calibration Check: 21 06/28/14 17:30 .896

Subject Name

Subject I.D.

Operator Name, I.D.
DEUTMAN 883

Location
CCPD BATVAN

AS IV Serial no: 897445
Version no: 5028

TEST RECORD 80841 s/
Temp Date Time 210L
VOIS: 191
12 06/28/14 17:00

Subject Name

Subject I.D.

Operator Name, I.D.
DEUTMAN 883

Location
CCPD BATVAN



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

MICHAEL DEUTMAN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 5000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 308.111 through 308.119 RSMo.

DATE 12/31/2013

NUMBER 230339

EXPIRES 12/31/2015

MO 660-0771 (6-10)

W. ...

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David Vasherby

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (PK-10)

