

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 12:16 pm, Sep 22, 2014

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097445 PRINTER SN 096.3580.951 DATE OF INSPECTION 09/19/2014

LOCATION OF INSTRUMENT (STREET AND CITY) 300 N.NEW BALLAS RD. CREVE COEUR TIME OF INSPECTION 5:37 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
PRINTER WORKING PROPERLY
TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
STANDARD SUPPLIER REPCO MARKETING INC. LOT # 13001 EXP. DATE 03/07/2015
SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2784 SIMULATOR EXP DATE 07/15/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .099% TEST 2 .098% TEST 3 .098%

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

Table with 7 columns: REFUSALS, (0-.04), (.05-.09), (.10-.14), (.15-.19), (OVER .19). Values: 0, 0, 0, 0, 0, 0.

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER
SIGNATURE: P.O. Michael T. Deutman 885
PRINT NAME: PO MICHAEL DEUTMAN #885
TYPE II PERMIT NUMBER/EXPIRATION DATE: 230339 12/31/2015
TELEPHONE NUMBER: (314) 432-8000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13001
EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl \pm .003 gms/dl w/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 \pm 3% gms/210L Breath when heated to 34 Degrees Celsius \pm 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

6S IV Serial no: 097445
Version no: 532B

TEST RECORD 00891 ^{sl}
Temp Date Time 2101

Air Blank:
09/19/14 17:37 .080
Calibration Check:
18 09/19/14 17:37 .099

Subject Name

Subject I.D.

Operator Name, I.D.
DEJMAN 883
Location
CCPD BATUAN

6S IV Serial no: 097445
Version no: 532B

TEST RECORD 00892 ^{sl}
Temp Date Time 2101

Air Blank:
09/19/14 17:38 .080
Calibration Check:
19 09/19/14 17:38 .098

Subject Name

Subject I.D.

Operator Name, I.D.
DEJMAN 883
Location
CCPD BATUAN

6S IV Serial no: 097445
Version no: 532B

TEST RECORD 00893 ^{sl}
Temp Date Time 2101

Air Blank:
09/19/14 17:40 .080
Calibration Check:
20 09/19/14 17:40 .098

Subject Name

Subject I.D.

Operator Name, I.D.
DEJMAN 883
Location
CCPD BATUAN

6S IV Serial no: 097445
Version no: 532B

TEST RECORD 00894 ^{sl}
Temp Date Time 2101

VOID: RFI
12 09/19/14 17:41

Subject Name

Subject I.D.

Operator Name, I.D.
DEJMAN 883
Location
CCPD BATUAN



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

MICHAEL DEUTMAN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 5000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/31/2013

NUMBER 230339

EXPIRES 12/31/2015

MO 680-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB 4 (16-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator: DEUTMAN, MICHAEL
Permit No: 230339
Date Issued: 12/31/2013 Date Expires: 12/31/2015