



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13290 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 31, 2013**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1202%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 29, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** \pm 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 097459
Version no: 004C

TEST RECORD 00142

Temp Date Time ^{s/} 210L

Air Blank:
11/26/14 15:19 .000
Calibration Check:
24 11/26/14 15:19 .099

Subject Name

Subject I.D.

Operator Name, I.D.

Levyne Carver CCI
Location
2920 N. Shamrock Rd
Jefferson City

AS IV Serial no: 097459
Version no: 004C

TEST RECORD 00143

Temp Date Time ^{s/} 210L

Air Blank:
11/26/14 15:21 .000
Calibration Check:
24 11/26/14 15:21 .099

Subject Name

Subject I.D.

Operator Name, I.D.

Levyne Carver CCI
Location
2920 N. Shamrock Rd.
Jefferson City

AS IV Serial no: 097459
Version no: 004C

TEST RECORD 00144

Temp Date Time ^{s/} 210L

Air Blank:
11/26/14 15:23 .000
Calibration Check:
24 11/26/14 15:23 .099

Subject Name

Subject I.D.

Operator Name, I.D.

Levyne Carver CCI
Location
2920 N. Shamrock Rd.
Jefferson City

AS IV Serial no: 097459
Version no: 004C

TEST RECORD 00145

Temp Date Time ^{s/} 210L

Void: RFI
12 11/26/14 15:24

Subject Name


Subject I.D.

Operator Name, I.D.

Levyne Carver CCI
Location
2920 N. Shamrock Rd
Jefferson City



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

DEWAYNE D CARVER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV W/PRINTER, DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 05/28/2013
NUMBER 230101
EXPIRES 05/28/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paul Vashelky
Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097459	PRINTER SN 096.3580.956	DATE OF INSPECTION 11/26/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 2920 North Shamrock Road, Jefferson City		TIME OF INSPECTION 3:26 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 13060 EXP. DATE 02/04/2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIMULATOR SN MP2122 SIMULATOR EXP DATE 06/03/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .039

TEST 2 .039

TEST 3 .039

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Serviced fuel cell diaphragm & locking assembly. Calibrated instrument.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Dewayne D. Carver
TYPE II PERMIT NUMBER/EXPIRATION DATE 230101 / 05/28/2015	TELEPHONE NUMBER (573) 751-4722

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13060** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **February 7, 2013**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.0482%** (w/vol) ethyl alcohol. The expiration date for this lot number is **February 4, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.040 g/210L** $\pm 3\%$.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN120110-04 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 097459
Version no: 004C

TEST RECORD 00146

Temp Date Time ^{s/} 210L

Air Blank:
11/26/14 15:26 .000
Calibration Check:
24 11/26/14 15:26 .039

Subject Name

Subject I.D.

Operator Name, I.D.

Leudwine Carver C/1
Location

2920 N. Shamrock Rd

Jefferson City

AS IV Serial no: 097459
Version no: 004C

TEST RECORD 00147

Temp Date Time ^{s/} 210L

Air Blank:
11/26/14 15:28 .000
Calibration Check:
25 11/26/14 15:28 .039

Subject Name

Subject I.D.

Operator Name, I.D.

Leudwine Carver C/1
Location

2920 N. Shamrock Rd

Jefferson City

AS IV Serial no: 097459
Version no: 004C

TEST RECORD 00148

Temp Date Time ^{s/} 210L

Air Blank:
11/26/14 15:29 .000
Calibration Check:
25 11/26/14 15:29 .039

Subject Name

Subject I.D.

Operator Name, I.D.

Leudwine Carver C/1
Location

2920 N. Shamrock Rd

Jefferson City

AS IV Serial no: 097459
Version no: 004C

TEST RECORD 00149

Temp Date Time ^{s/} 210L

Void: RFI
12 11/26/14 15:31

Subject Name

Subject I.D.

Operator Name, I.D.


Leudwine Carver C/1
Location

2920 N. Shamrock Rd

Jefferson City



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

DEWAYNE D CARVER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV W/PRINTER, DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 05/28/2013
 NUMBER 230101
 EXPIRES 05/28/2015

W. S. ...

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David VanDerKam

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

MO 560-0771 (6-10)