MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED REPORT #7
By Carol Day at 3:13 pm, Jan 24, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
094804	097.3584.334	01/17/2014
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
201 W. Second Street, Sedalia, MO 65301		10:13 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- ☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- ☑ PRINTER WORKING PROPERLY
- ☑ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- ☑ SIMULATOR SOLUTION ☐ COMPRESSED ETHANOL-GAS MIXTURE
- ☑ STANDARD SUPPLIER RepCo LOT # 13002 EXP. DATE 06/19/2015
- ☑ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD 2306 SIMULATOR EXP DATE 10/17/2014
- ☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
  Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛ .100	TEST 2 ☛ .100	TEST 3 ☛ .100

- ☑ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
0	0	0	0	1	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

This instrument was calibrated. This instrument conforms to the Department of Health standards.

INSPECTING OFFICER

SIGNATURE	PRINT NAME
▸ Victoria A. Kottman	Victoria A. Kottman
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
230120 06-12-2015	(660) 826-8100

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

REPCO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13002
EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number **13002** of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain **.1217** gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of **.100** +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is **June 20, 2013** The expiration date for this lot number is **June 19, 2015** at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President
RepCo Marketing, Inc.

Form RM 02

AS IV Serial no: 094804
Version no: 532B

TEST RECORD 00142

Temp	Date	Time	g/ 210L
Air Blank:			
	01/17/14	10:13	.000
Calibration:			
23	01/17/14	10:13	.100

Subject Name

Subject I.D.
Victoria Kottman
Operator Name, I.D.
230120 6-12-15
Location
201 W. Second
Sedalia MO
65301

AS IV Serial no: 094804
Version no: 532B

TEST RECORD 00144

Temp	Date	Time	g/ 210L
Air Blank:			
	01/17/14	10:20	.000
Calibration Check:			
24	01/17/14	10:20	.100

Subject Name

Subject I.D.
Victoria Kottman
Operator Name, I.D.
230120 6-12-15
Location
201 W. Second Str
Sedalia MO
65301

AS IV Serial no: 094804
Version no: 532B

TEST RECORD 00145

Temp	Date	Time	g/ 210L
Air Blank:			
	01/17/14	10:22	.000
Calibration Check:			
24	01/17/14	10:22	.100

Subject Name

Subject I.D.
Victoria Kottman
Operator Name, I.D.
230120 6-12-15
Location
201 W. Second
Sedalia MO
65301

AS IV Serial no: 094804
Version no: 532B

TEST RECORD 00146

Temp	Date	Time	g/ 210L
Air Blank:			
	01/17/14	10:24	.000
Calibration Check:			
25	01/17/14	10:24	.100

Subject Name

Subject I.D.
Victoria Kottman
Operator Name, I.D.
230120 6-12-15
Location
201 W. Second
Sedalia MO
65301

AS IV Serial no: 094804
Version no: 532B

TEST RECORD 00149

Temp	Date	Time	g/ 210L
VOID: RFI			
12	01/17/14	10:30	

Subject Name
Test A
Subject I.D.
Victoria Kottman
Operator Name, I.D.
230120 6-12-15
Location
201 W. Second
Sedalia MO

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

VICTORIA A KOTTMAN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 06/12/2013

NUMBER 230120

EXPIRES 06/12/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)