

PERMIT
TYPE II

JENNIFER L STEPHENS

is hereby authorized to instruct and supervise operators in the use of the following breath analyzers:

ALCO-SENSOR IV WITH PUNTER. DATAMASTER
for the determination of the alcoholic content of blood from a sample of expired air. Permits issued under the provisions of sections 577.020 through 577.041 RSMo and 306.111 through 306.119 RSMo

DATE 8/14/2013

NUMBER 230170

EXPIRES 8/14/2015

O.S.M.S. 0771 1E 1U1

Wesley
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Bob Verduy
acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAE- PRE 1U1

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD

The permit cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator: STEPHENS, JENNIFER
Permit No. 230170
Date issued 8/14/2013 Date Expires 8/14/2015



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13280** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2013**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1217%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 16, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 099360
Version no: 532B

TEST RECORD 00251

Temp Date Time 210L ^{s/}

Air Blank:
07/04/14 14:10 .000
Calibration Check:
24 07/04/14 14:10 .096
TEST # 2

Subject Name

Subject I.I.

STEPHENS 099

Operator Name, I.I.

RPP

Location

AS IV Serial no: 099360
Version no: 532B

TEST RECORD 00250

Temp Date Time 210L ^{s/}

Air Blank:
07/04/14 14:09 .000
Calibration Check:
23 07/04/14 14:09 .096
TEST # 1

Subject Name

Subject I.I.

STEPHENS 099

Operator Name, I.I.

RPP

Location

AS IV Serial no: 099360
Version no: 532B

TEST RECORD 00252 a/
Temp Date Time 210L

Air Blank: 07/04/14 14:12 .000
Calibration Check: 25 02/04/14 14:12 .096
TEST # 3

Subject Name

Subject I.D.

STEPHEN C.F.

Operator Name, I.D.

RCO

Location

AS IV Serial no: 099360
Version no: 532B

TEST RECORD 00253 a/
Temp Date Time 210L

UNIT: RFI 12 07/04/14 14:13
RFI TEST

Subject Name

Subject I.D.

STEPHEN C.F.

Operator Name, I.D.

RCO

Location