



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

**RECEIVED**  
 By Carol Day at 10:04 am, Dec 10, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance of the instrument. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 099364	PRINTER SN 097.3584.339	DATE OF INSPECTION 11/24/2014
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Franklin County Sheriffs Office, #1 Bruns Ln. Union MO 63084	TIME OF INSPECTION 4:14 pm
--	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER INTOXIMETERS LOT # AG423201 EXP. DATE 08/20/2016
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIMULATOR SN \_\_\_\_\_ SIMULATOR EXP DATE \_\_\_\_\_

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .102	TEST 2 ➔ .102	TEST 3 ➔ .101
---------------	---------------	---------------

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT NAME Cpl. M. Richardson 1271
TYPE II PERMIT NUMBER/EXPIRATION DATE 240127 04/03/2016	TELEPHONE NUMBER (636) 583-2560

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901

AS IV Serial no: 099364  
Version no: 532B

TEST RECORD 00262

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
11/24/14 16:14 .000  
Calibration Check:  
22 11/24/14 16:14 .102

Subject Name

Subject I.D.

  
Operator Name, I.D.

Location

AS IV Serial no: 099364  
Version no: 532B


TEST RECORD 00263

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
11/24/14 16:16 .000  
Calibration Check:  
22 11/24/14 16:16 .102

Subject Name

Subject I.D.

  
Operator Name, I.D.

Location

AS IV Serial no: 099364  
Version no: 532B


TEST RECORD 00264

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
11/24/14 16:17 .000  
Calibration Check:  
23 11/24/14 16:17 .101

Subject Name

Subject I.D.

  
Operator Name, I.D.

Location

AS IV Serial no: 099364  
Version no: 532B

TEST RECORD 00265

Temp Date Time <sup>g/</sup> 210L

VOID: RFI  
12 11/24/14 16:18

Subject Name

Subject I.D.

  
Operator Name, I.D.

Location



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

**Certificate of Analysis**

**Customer Name**  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 20-Aug-2014

**Lot #** AG423201

<b><u>Exp. Date</u></b> 20-Aug-2016	<b><u>Cyl. Type</u></b> 108	<b><u>Component</u></b> Ethanol Nitrogen	<b><u>Certified Concentration</u></b> 0.100 ± 2% BrAC (272 ppm) Balance
--	--------------------------------	--	---

Certification Traceable to N.I.S.T. RGM Ethanol Standards:


<b><u>Serial No.</u></b>	<b><u>Concentration</u></b>	<b><u>Serial No.</u></b>	<b><u>Concentration</u></b>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
 Date: 2014.08.22 12:04:20 -05:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Analyst:   
 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**

**MICHAEL RICHARDSON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 806.111 through 806.119 RSMo.

DATE 4/3/2014

NUMBER 240127

EXPIRES 4/3/2016

MO 690.0171 (6-10)

*W. W. ...*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Gail Vosterly*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (6-10)