



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

received 1/9/14-cd

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

**REVIEWED**  
By Carol Day at 8:08 am, Feb 04, 2014

ALCO SENSOR IV SN 100284	PRINTER SN 093.3563.011	DATE OF INSPECTION 01/03/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 777 Broadway Hannibal MO		TIME OF INSPECTION 11:20 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Repco LOT # 13002 EXP. DATE 06/19/2015
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2224 SIMULATOR EXP DATE 08/09/2014

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .101	TEST 2  .100	TEST 3  .100
--------------	--------------	--------------

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	1	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	3	(OVER .19)	2
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME Jacob Nacke
TYPE II PERMIT NUMBER/EXPIRATION DATE 220111 / 05/09/2014	TELEPHONE NUMBER (573) 221-0987

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901

AS IV Serial no: 100284  
Version no: 004C

TEST RECORD 00554

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
01/03/14 11:52 .000  
Calibration Check:  
18 01/03/14 11:52 .101

Subject Name

Subject I.D.

Operator Name, I.D.

Location

*J. M. M.*  
*220111 / 5/9/14*

2

AS IV Serial no: 100284  
Version no: 004C

TEST RECORD 00555

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
01/03/14 11:54 .000  
Calibration Check:  
19 01/03/14 11:54 .100

Subject Name

Subject I.D.

Operator Name, I.D.

*J. M. M.*  
Location  
*220111 / 5/9/14*

3

AS IV Serial no: 100284  
Version no: 004C

TEST RECORD 00556

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
01/03/14 11:56 .000  
Calibration Check:  
19 01/03/14 11:56 .100

Subject Name

Subject I.D.

Operator Name, I.D.

*J. M. M.*  
Location  
*220111 / 5/9/14*

RFI

AS IV Serial no: 100284  
Version no: 004C

TEST RECORD 00557

Temp Date Time 210L <sup>s/</sup>

Void: RFI  
12 01/03/14 11:58

Subject Name

Subject I.D.

Operator Name, I.D.

*J. M. M.*  
Location  
*220111 / 5/9/14*

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 13002**

**EXPIRATION DATE: June 19, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013  
The expiration date for this lot number is June 19, 2015 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

State of Missouri  
DEPARTMENT OF HEALTH



P E R M I T  
TYPE II



JACOB W NACKE

---

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV W/PRINTER**

---

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 05/09/2012

Number **220111**

Expires 05/09/2014

Director of State Public Health Laboratory

Director, Department of Health