

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



J ALEX GROTE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 05/09/2012

Number 220106

Expires 05/09/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)

AS IV Serial no: 100204
Version no: 004C

TEST RECORD - RPRINT

TEST RECORD 00632

Temp Date Time 210L

Air Blank:
05/07/14 13:45 .000
Calibration Check:
23 05/07/14 13:45 .099

Subject Name

Subject I.D.

Operator Name, I.D.

Grote 21

Location

220106 5-9-14

AS IV Serial no: 100204
Version no: 004C

TEST RECORD 00634

Temp Date Time 210L

Air Blank:
05/07/14 13:53 .000
Calibration Check:
24 05/07/14 13:53 .099

Subject Name

Subject I.D.

Operator Name, I.D.

Brote 21

Location

220106 5-9-14

AS IV Serial no: 100204
Version no: 004C

TEST RECORD 00633

Temp Date Time 210L

Air Blank:
05/07/14 13:51 .000
Calibration Check:
24 05/07/14 13:51 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Grote 21

Location

220106 5-9-14

AS IV Serial no: 100204
Version no: 004C

TEST RECORD 00635

Temp Date Time 210L

Void: RFI
12 05/07/14 13:54

Subject Name

Subject I.D.

Operator Name, I.D.

Grote 21

Location

220106 5-9-14