MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

# ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

**RECEIVED**
By Carol Day at 11:16 am, Nov 07, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| ALCO SENSOR IV SN | PRINTER SN | DATE OF INSPECTION |
|---|---|---|
| 100284 | 096.3580.945 | 11/05/2014 |
| LOCATION OF INSTRUMENT (STREET AND CITY) | | TIME OF INSPECTION |
| 777 Broadway Hannibal | | 2:44 pm |

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- ☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- ☑ PRINTER WORKING PROPERLY
- ☑ TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

☑ SIMULATOR SOLUTION ☐ COMPRESSED ETHANOL-GAS MIXTURE

☑ STANDARD SUPPLIER Repco LOT # 14001 EXP. DATE 04/30/2016

☑ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN SD2224 SIMULATOR EXP DATE 08/22/2015

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

| TEST 1 ☛ .098 | TEST 2 ☛ .098 | TEST 3 ☛ .098 |
|---|---|---|

☑ RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

| REFUSALS | (0-.04) | (.05-.09) | (.10-.14) | (.15-.19) | (OVER .19) |
|---|---|---|---|---|---|
| 0 | 0 | 0 | 2 | 2 | 0 |

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Changed time.

**INSPECTING OFFICER**

| SIGNATURE | PRINT NAME |
|---|---|
| ▸ 36 | Jacob Nacke |
| TYPE II PERMIT NUMBER/EXPIRATION DATE | TELEPHONE NUMBER |
| 240225 / 05/09/2016 | (573) 221-0987 |

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901


AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

AS IV Serial no: 100284
Version no: 004C

TEST RECORD 00819

| Temp | Date | Time | 210L |
|---|---|---|---|
| Air Blank: | | | |
| | 11/05/14 | 14:44 | .000 |
| Calibration Check: | | | |
| 24 | 11/05/14 | 14:44 | .098 |

Subject Name

Subject I.D.

Operator Name, I.D.
NACKE #36

Location
240225 / 5/9/16

)

AS IV Serial no: 100284
Version no: 004C

TEST RECORD 00820

| Temp | Date | Time | 210L |
|---|---|---|---|
| Air Blank: | | | |
| | 11/05/14 | 14:46 | .000 |
| Calibration Check: | | | |
| 24 | 11/05/14 | 14:46 | .098 |

Subject Name

Subject

Operator Name, I.D.
NACKE #36

Location
240225 5/9/16

②

AS IV Serial no: 100284
Version no: 004C

TEST RECORD 00821

| Temp | Date | Time | 210L |
|---|---|---|---|
| Air Blank: | | | |
| | 11/05/14 | 14:47 | .000 |
| Calibration Check: | | | |
| 25 | 11/05/14 | 14:47 | .098 |

Subject Name

Subject I.D.

Operator Name
NACKE #36

Location
240225 5/9/16

③

AS IV Serial no: 100284
Version no: 004C

TEST RECORD 00822

| Temp | Date | Time | 210L |
|---|---|---|---|
| Void: RFI | | | |
| 12 | 11/05/14 | 14:49 | |

Subject Name

Subject

Operator Name, I.D.
NACKE #36

Location
240225 5/9/16

RFI

*REPCO MARKETING INC.*

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

# CERTIFICATE OF ANALYSIS

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**
**LOT NUMBER: 14001**
**EXPIRATION DATE: April 30, 2016** at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number **14001** of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain **.1216** gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of **.100** +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is **May 1, 2014** The expiration date for this lot number is **April 30, 2016** at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President
RepCo Marketing, Inc.

Form RM 02

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

# PERMIT

## TYPE II

## JACOB W NACKE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

## ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/9/2014

NUMBER 240225

EXPIRES 5/9/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** NACKE, JACOB

**Permit No** 240225

**Date Issued** 5/9/2014 **Date Expires** 5/9/2016