MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 1:20 pm, Jan 16, 2014

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
100286	093.3563.13	01/15/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 7447 Dale Avenue, Richmond Heights, MO 63117		TIME OF INSPECTION 12:04 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- ☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- ☑ PRINTER WORKING PROPERLY
- ☑ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☐ SIMULATOR SOLUTION ☑ COMPRESSED ETHANOL-GAS MIXTURE

☑ STANDARD SUPPLIER Airgas LOT # AG302902 EXP. DATE 09/29/2014

☐ SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☐ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- ☑ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛ 0.080%	TEST 2 ☛ 0.080%	TEST 3 ☛ 0.079%

☑ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	1	(OVER .19)	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE	PRINT NAME
▸ Sgt. Stewart Deken 151	Sgt. Stewart Deken, 151
TYPE II PERMIT NUMBER/EXPIRATION DATE 220402 12/20/2014	TELEPHONE NUMBER (314) 647-5656

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

Airgas

Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 30-Jan-2013

Lot # AG302902

Exp. Date	Cyl. Type	Component	Certified Concentration
29-Sep-2014	30	Ethanol	0.052 ± 0.002 BrAC (223 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

Serial No.	Concentration	Serial No.	Concentration
EB0010[illegible]	391.[illegible] ppm	[illegible]	399.[illegible] ppm
EB0010570	258.4 ppm	[illegible]	258.3 ppm
EB0010[illegible]	208.[illegible] ppm	[illegible]	209.2 ppm
EB0010[illegible]	101.[illegible] ppm	[illegible]	104.9 ppm
EB0010[illegible]	53.0 ppm	[illegible]	52.4 ppm

Analytical Method: NDIR

Analyst: [signature]

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

AS IV Serial no: 100286
Version no: 004C

TEST RECORD 00683

Temp	Date	Time	210L
Air Blank:			
	01/15/14	12:04	.000
Subject Test: Man			
20	01/15/14	12:04	.080

Subject Name
Test

Subject I.D.

Operator Name, I.D.
S. Styl 101

Location
Richmond Heights PD

AS IV Serial no: 100286
Version no: 004C

TEST RECORD 00684

Temp	Date	Time	210L
Air Blank:			
	01/15/14	12:06	.000
Subject Test: Man			
20	01/15/14	12:06	.080

Subject Name
Test

Subject I.D.

Operator Name, I.D.
S. Styl 101

Location
Richmond Heights PD

AS IV Serial no: 100286
Version no: 004C

TEST RECORD 00685

Temp	Date	Time	210L
Air Blank:			
	01/15/14	12:07	.000
Subject Test: Man.			
20	01/15/14	12:07	.079

Subject Name
Test

Subject I.D.

Operator Name, I.D.
S. Styl 101

Location
Richmond Heights PD

AS IV Serial no: 100286
Version no: 004C

TEST RECORD 00686

Temp	Date	Time	210L
Void: RFI			
12	01/15/14	12:08	

Subject Name
RFI Test

Subject I.D.

Operator Name, I.D.
S. Styl 101

Location
Richmond Heights PD

State of Missouri
DEPARTMENT OF HEALTH

PERMIT
TYPE II

STEWART J DEKEN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 12/20/2012

Number 220402

Expires 12/20/2014

Director of State Public Health Laboratory

ACTING DIRECTOR

Director, Department of Health

MO 580-0771 (7-88)

Lab. 4 (87-88)