MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 12:22 pm, Apr 14, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
100292	093.3563.020	04/03/2014

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
104 W. School, Hamilton	8:25 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- ☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- ☑ PRINTER WORKING PROPERLY
- ☑ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☑ SIMULATOR SOLUTION ☐ COMPRESSED ETHANOL-GAS MIXTURE

☑ STANDARD SUPPLIER Guth Laboratories LOT # 13210 EXP. DATE 07/29/2015

☑ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD3147 SIMULATOR EXP DATE 08/27/2014

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
.101	.101	.101

☑ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument performing within standards.

INSPECTING OFFICER

SIGNATURE	PRINT NAME
Sgt. Kimberly A. Grebner	Sgt. Kimberly A. Grebner
TYPE INPERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
220190 08/13/2014	(816) 583-7311

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

GUTH LABORATORIES, INC.
690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13210 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **July 31, 2013,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1216%** (w/vol) ethyl alcohol. The expiration date for this lot number is **July 29, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Paulcy, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FNI22211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 100292
Version no: 004C

TEST RECORD 00466

Temp	Date	Time	s/ 210L
Air Blank:			
	04/03/14	08:25	.000
Calibration Check:			
23	04/03/14	08:25	.101

Subject Name

Subject I.D.

Operator Name, I.D.
Lt Kimberly A. Grebner
Location #310

AS IV Serial no: 100292
Version no: 004C

TEST RECORD 00467

Temp	Date	Time	s/ 210L
Air Blank:			
	04/03/14	08:27	.000
Calibration Check:			
24	04/03/14	08:27	.101

Subject Name

Subject I.D.

Operator Name, I.D.
Lt Kimberly A. Grebner #310
Location

AS IV Serial no: 100292
Version no: 004C

TEST RECORD 00468

Temp	Date	Time	s/ 210L
Air Blank:			
	04/03/14	08:29	.000
Calibration Check:			
25	04/03/14	08:29	.101

Subject Name

Subject I.D.

Operator Name, I.D.
Lt Kimberly A. Grebner
Location #310

AS IV Serial no: 100292
Version no: 004C

TEST RECORD 00469

Temp	Date	Time	s/ 210L
Void: RFI			
12	04/03/14	08:30	

Subject Name

Subject I.D.

Operator Name, I.D.
Lt Kimberly A. Grebner
Location #310

State of Missouri
DEPARTMENT OF HEALTH

PERMIT
TYPE II

KIMBERLY A GREBNER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 08/13/2012

Number 220190

Expires 08/13/2014

Director of State Public Health Laboratory

Director, Department of Health

MO 530-0771 (7-86)

Lab. 4 (R7-88)