MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

# ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 9:43 am, Aug 06, 201

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| ALCO SENSOR IV SN | PRINTER SN | DATE OF INSPECTION |
|---|---|---|
| 102458 | 093.3578.252 | 08/05/2014 |
| LOCATION OF INSTRUMENT (STREET AND CITY) | | TIME OF INSPECTION |
| 600 East Walnut Columbia, Missouri | | 10:07 pm |

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- [x] TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- [x] PRINTER WORKING PROPERLY
- [x] TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- [ ] SIMULATOR SOLUTION
- [x] COMPRESSED ETHANOL-GAS MIXTURE

- [x] STANDARD SUPPLIER Intoximeters LOT # AG414902 EXP. DATE 05/29/2016

- [ ] SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

- [x] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - [ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

| TEST 1 ☛ .099 | TEST 2 ☛ .099 | TEST 3 ☛ .099 |
|---|---|---|

- [x] RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

| REFUSALS | (0-.04) | (.05-.09) | (.10-.14) | (.15-.19) | (OVER .19) |
|---|---|---|---|---|---|
| 2 | 1 | 0 | 4 | 2 | 2 |

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

| SIGNATURE | PRINT NAME |
|---|---|
| ▸ | Nathan Turner |
| TYPE II PERMIT NUMBER/EXPIRATION DATE | TELEPHONE NUMBER |
| 220411 12/20/2014 | (573) 874-7652 |

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff MO 63901

Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

# Certificate of Analysis

**Customer Name**
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

**Test Date:** 2-Jun-2014

## Lot # AG414902

| Exp. Date | Cyl. Type | Component | Certified Concentration |
|---|---|---|---|
| 29-May-2016 | 55 | Ethanol | 0.100 ± 2% BrAC (272 ppm) |
| | | Nitrogen | Balance |

**Certification Traceable to N.I.S.T. RGM Ethanol Standards:**

| Serial No. | Concentration | Serial No. | Concentration |
|---|---|---|---|
| EB0010581 | 391.8 ppm | EB0010603 | 392.5 ppm |
| EB0010570 | 259.8 ppm | EB0010559 | 258.9 ppm |
| EB0010285 | 209.0 ppm | EB0010595 | 208.9 ppm |
| EB0010561 | 103.7 ppm | EB0010562 | 104.9 ppm |
| EB0010681 | 52.22 ppm | EB0010579 | 52.94 ppm |

**Analytical Method:** NDIR

Digitally signed by Quality Control
Date: 2014.06.03 09:10:02 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

**Analyst:** Rod Marsala

*ISO 17025:2005 A2LA accredited. Certificate Number 2989.01*

AS IV Serial no: 102458
Version no: 004C

TEST RECORD 00966

| Temp | Date | Time | 9/ 210L |
|---|---|---|---|

Air Blank:
08/05/14 22:11 .000
Calibration Check:
26 08/05/14 22:11 .099

Subject Name
Test

Subject I.D.

Operator Name, I.D.
Turner 1985

Location
CPD

AS IV Serial no: 102458
Version no: 004C

TEST RECORD 00964

| Temp | Date | Time | 9/ 210L |
|---|---|---|---|

Air Blank:
08/05/14 22:07 .000
Calibration Check:
24 08/05/14 22:07 .099

Subject Name
Test

Subject I.D.

Operator Name, I.D.
Turner 1985

Location
CPD

AS IV Serial no: 102458
Version no: 004C

TEST RECORD 00967

| Temp | Date | Time | 9/ 210L |
|---|---|---|---|

Void: RFI
12 08/05/14 22:12

Subject Name
Test

Subject I.D.

Operator Name, I.D.
Turner 1985

Location
CPD

AS IV Serial no: 102458
Version no: 004C

TEST RECORD 00965

| Temp | Date | Time | 9/ 210L |
|---|---|---|---|

Air Blank:
08/05/14 22:09 .000
Calibration Check:
25 08/05/14 22:09 .099

Subject Name
Test

Subject I.D.

Operator Name, I.D.
Turner 1985

Location
CPD