



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

RECEIVED 2/4/14 #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and Send copy to Department of Health and Senior Services; retain original in department file.

**REVIEWED**  
 By Carol Day at 9:26 am, Mar 14, 2014

ALCO SENSOR IV SN <i>102461</i>	PRINTER SN <i>096.3580.922</i>	DATE OF INSPECTION <i>31 Jan 2014</i>
LOCATION OF INSTRUMENT (STREET AND CITY) <i>106 Progress Drive Sullivan MO</i>		TIME OF INSPECTION

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C) *26°C*
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER *Cuth Labs* LOT # *13210* EXP. DATE *07/29/15*
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) *34°C* SIMULATOR SN *S03322* SIMULATOR EXP DATE *01/25/15*

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • <i>.097%</i>	TEST 2 • <i>.098%</i>	TEST 3 • <i>.097%</i>
-----------------------	-----------------------	-----------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <i>0</i>	(0-.04) <i>1</i>	(.05-.09) <i>1</i>	(.10-.14) <i>0</i>	(.15-.19) <i>2</i>	(OVER .19) <i>1</i>
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).  
*Instrument is operating within D.O.H. specifications*

**INSPECTING OFFICER**

SIGNATURE <i>Gregory A. West</i>	PRINT NAME <i>Gregory A. West</i>
TYPE II PERMIT NUMBER/EXPIRATION DATE <i>230109 06/04/15</i>	TELEPHONE NUMBER <i>573-468-8001</i>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901



**GUTH LABORATORIES, INC.**

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13210 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1216% (w/vol) ethyl alcohol. The expiration date for this lot number is July 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FNI22211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

AS IV Serial no: 182461  
Version no: 004C

TEST RECORD 00257 s/

Temp Date Time 210L

Void: RFI  
12 01/31/14 11:36

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 182461  
Version no: 004C

TEST RECORD 00256 s/

Temp Date Time 210L

Air Blank:  
01/31/14 11:33 .000

Calibration Check:  
25 01/31/14 11:33 .097

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 182461  
Version no: 004C

TEST RECORD 00255 s/

Temp Date Time 210L

Air Blank:  
01/31/14 11:32 .000

Calibration Check:  
26 01/31/14 11:32 .098

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 182461  
Version no: 004C

TEST RECORD 00254 s/

Temp Date Time 210L

Air Blank:  
01/31/14 11:30 .000

Calibration Check:  
26 01/31/14 11:30 .097

Subject Name

Subject I.D.

Operator Name, I.D.

Location



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**

**GREGORY A WEST**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 06/04/2013

NUMBER 230109

EXPIRES 06/04/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Michael Vickard*  
 Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES