



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED 6/14/14-CD

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check. Send copy to Department of Health and Senior Services; retain original in department file.

REVIEWED instrument is repaired
By Carol Day at 1:51 pm, Jun 26, 2014

ALCO SENSOR IV SN <u>102466</u>	PRINTER SN <u>079.3418.057</u>	DATE OF INSPECTION <u>06/04/14</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>106 PROGRESS SULLIVAN MO 63080</u>		TIME OF INSPECTION <u>0542</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C) 24°C
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER GUTH LOT # 14030 EXP. DATE 01/20/16
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN 503322 SIMULATOR EXP DATE 01/28/15

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 <u>0.099</u>	TEST 2 <u>0.099</u>	TEST 3 <u>.100</u>
---------------------	---------------------	--------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <u>2</u>	(0-.04) <u>2</u>	(.05-.09) <u>0</u>	(.10-.14) <u>0</u>	(.15-.19) <u>ONE</u>	(OVER .19) <u>3</u>
-------------------	------------------	--------------------	--------------------	----------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSTRUMENT IS OPERATING WITHIN D.O.H. SPECIFICATIONS

INSPECTING OFFICER

SIGNATURE <u>Gregory A. West</u>	PRINT NAME <u>Gregory A. West</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>230109 06/04/15</u>	TELEPHONE NUMBER <u>573-468-8001</u>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14030 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 22, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1215% (w/vol) ethyl alcohol. The expiration date for this lot number is January 20, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.190 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IU Serial no: 102466
Version no: 5328

TEST RECORD 00036

Temp Date Time 210L s/

Air Blank: 05/04/14 05:42 .000

Calibration Check: 24 06/04/14 05:42 .099

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IU Serial no: 102466
Version no: 5328

TEST RECORD 00037

Temp Date Time 210L s/

Air Blank: 06/04/14 05:46 .000

Calibration Check: 25 06/04/14 05:46 .099

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IU Serial no: 102466
Version no: 5328

TEST RECORD 00038

Temp Date Time 210L s/

Air Blank: 06/04/14 05:48 .000

Calibration Check: 25 06/04/14 05:48 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IU Serial no: 102466
Version no: 5328

TEST RECORD 00039

Temp Date Time 210L s/

VOID: RFI 12 06/04/14 05:50

Subject Name

Subject I.D.

Operator Name, I.D.

Location



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

GREGORY A WEST

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 06/04/2013

NUMBER 230109

EXPIRES 06/04/2015

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (RS-10)